

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX061022\
 Data File : VX029362.D
 Acq On : 10 Jun 2022 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 00:51:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	116	0.00
2 T	Dichlorodifluoromethane	50.000	45.837	8.3	100	0.00
3 P	Chloromethane	50.000	41.661	16.7	97	0.00
4 C	Vinyl Chloride	50.000	44.119	11.8#	96	0.00
5 T	Bromomethane	50.000	44.625	10.8	103	0.00
6 T	Chloroethane	50.000	55.956	-11.9	115	0.00
7 T	Trichlorofluoromethane	50.000	45.736	8.5	100	0.00
8 T	Diethyl Ether	50.000	46.460	7.1	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.138	7.7	104	0.00
10 T	Methyl Iodide	50.000	43.571	12.9	95	0.00
11 T	Tert butyl alcohol	250.000	176.808	29.3#	80	-0.01
12 CM	1,1-Dichloroethene	50.000	44.459	11.1#	98	0.00
13 T	Acrolein	250.000	190.245	23.9	89	0.00
14 T	Allyl chloride	50.000	45.059	9.9	99	0.00
15 T	Acrylonitrile	250.000	209.363	16.3	91	0.00
16 T	Acetone	250.000	243.524	2.6	111	0.00
17 T	Carbon Disulfide	50.000	43.410	13.2	95	0.00
18 T	Methyl Acetate	50.000	41.382	17.2	93	0.00
19 T	Methyl tert-butyl Ether	50.000	45.654	8.7	99	0.00
20 T	Methylene Chloride	50.000	42.164	15.7	97	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.207	11.6	99	0.00
22 T	Diisopropyl ether	50.000	46.536	6.9	99	0.00
23 T	Vinyl Acetate	250.000	236.477	5.4	97	0.00
24 P	1,1-Dichloroethane	50.000	44.645	10.7	97	0.00
25 T	2-Butanone	250.000	222.885	10.8	96	0.00
26 T	2,2-Dichloropropane	50.000	51.258	-2.5	112	0.00
27 T	cis-1,2-Dichloroethene	50.000	44.090	11.8	100	0.00
28 T	Bromochloromethane	50.000	45.334	9.3	100	0.00
29 T	Tetrahydrofuran	250.000	213.861	14.5	91	-0.01
30 C	Chloroform	50.000	45.439	9.1#	100	0.00
31 T	Cyclohexane	50.000	44.890	10.2	99	0.00
32 T	1,1,1-Trichloroethane	50.000	46.463	7.1	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	43.364	13.3	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	113	0.00
35 S	Dibromofluoromethane	50.000	42.112	15.8	102	0.00
36 T	1,1-Dichloropropene	50.000	46.705	6.6	100	0.00
37 T	Ethyl Acetate	50.000	45.272	9.5	93	0.00
38 T	Carbon Tetrachloride	50.000	47.857	4.3	102	0.00
39 T	Methylcyclohexane	50.000	48.149	3.7	102	0.00
40 TM	Benzene	50.000	45.784	8.4	98	0.00
41 T	Methacrylonitrile	50.000	45.967	8.1	94	0.00
42 TM	1,2-Dichloroethane	50.000	47.089	5.8	100	0.00
43 T	Isopropyl Acetate	50.000	47.256	5.5	98	0.00
44 TM	Trichloroethene	50.000	46.093	7.8	101	0.00
45 C	1,2-Dichloropropane	50.000	45.711	8.6#	98	0.00
46 T	Dibromomethane	50.000	47.073	5.9	100	0.00
47 T	Bromodichloromethane	50.000	49.145	1.7	102	0.00
48 T	Methyl methacrylate	50.000	46.363	7.3	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061022\
 Data File : VX029362.D
 Acq On : 10 Jun 2022 10:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 11 00:51:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 08 03:00:38 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	783.564	21.6	80	0.00
50 S	Toluene-d8	50.000	41.191	17.6	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	231.803	7.3	94	0.00
52 CM	Toluene	50.000	46.558	6.9#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	48.668	2.7	107	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.538	-1.1	101	0.00
55 T	1,1,2-Trichloroethane	50.000	46.445	7.1	98	0.00
56 T	Ethyl methacrylate	50.000	49.254	1.5	97	0.00
57 T	1,3-Dichloropropane	50.000	45.729	8.5	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	223.858	10.5	92	0.00
59 T	2-Hexanone	250.000	235.640	5.7	94	0.00
60 T	Dibromochloromethane	50.000	51.226	-2.5	104	0.00
61 T	1,2-Dibromoethane	50.000	46.362	7.3	97	0.00
62 S	4-Bromofluorobenzene	50.000	44.582	10.8	107	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	111	0.00
64 T	Tetrachloroethene	50.000	49.041	1.9	104	0.00
65 PM	Chlorobenzene	50.000	46.221	7.6	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.188	-0.4	101	0.00
67 C	Ethyl Benzene	50.000	48.400	3.2#	99	0.00
68 T	m/p-Xylenes	100.000	98.138	1.9	99	0.00
69 T	o-Xylene	50.000	49.423	1.2	100	0.00
70 T	Styrene	50.000	50.025	-0.0	100	0.00
71 P	Bromoform	50.000	53.603	-7.2	105	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	113	0.00
73 T	Isopropylbenzene	50.000	47.964	4.1	101	0.00
74 T	N-amyl acetate	50.000	48.964	2.1	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	44.128	11.7	96	0.00
76 T	1,2,3-Trichloropropane	50.000	46.739	6.5	98	0.00
77 T	Bromobenzene	50.000	46.535	6.9	101	0.00
78 T	n-propylbenzene	50.000	49.071	1.9	102	0.00
79 T	2-Chlorotoluene	50.000	48.230	3.5	103	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.438	3.1	102	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.401	5.2	102	0.00
82 T	4-Chlorotoluene	50.000	48.183	3.6	102	0.00
83 T	tert-Butylbenzene	50.000	48.416	3.2	101	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.271	1.5	103	0.00
85 T	sec-Butylbenzene	50.000	49.814	0.4	103	0.00
86 T	p-Isopropyltoluene	50.000	50.235	-0.5	104	0.00
87 T	1,3-Dichlorobenzene	50.000	47.738	4.5	103	0.00
88 T	1,4-Dichlorobenzene	50.000	46.466	7.1	102	0.00
89 T	n-Butylbenzene	50.000	52.202	-4.4	107	0.00
90 T	Hexachloroethane	50.000	51.564	-3.1	109	0.00
91 T	1,2-Dichlorobenzene	50.000	46.246	7.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.304	5.4	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.987	2.0	103	0.00
94 T	Hexachlorobutadiene	50.000	47.774	4.5	105	0.00
95 T	Naphthalene	50.000	48.857	2.3	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061022\
Data File : VX029362.D
Acq On : 10 Jun 2022 10:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 11 00:51:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060722W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 08 03:00:38 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.868	4.3	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6