

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061024\
 Data File : VX041798.D
 Acq On : 10 Jun 2024 10:25
 Operator : JC/MD
 Sample : VX0610MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0610MBL01

Quant Time: Jun 11 03:43:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	152140	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	239113	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242255	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	92123	49.371	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.740%
35) Dibromofluoromethane	5.379	113	75946	45.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.880%
50) Toluene-d8	8.647	98	292012	50.551	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.100%
62) 4-Bromofluorobenzene	11.079	95	112736	52.878	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.760%
Target Compounds						
16) Acetone	2.386	43	6113	4.772	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061024\
Data File : VX041798.D
Acq On : 10 Jun 2024 10:25
Operator : JC/MD
Sample : VX0610MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0610MBL01

Quant Time: Jun 11 03:43:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

