

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061024\
 Data File : VX041809.D
 Acq On : 10 Jun 2024 14:44
 Operator : JC/MD
 Sample : P2767-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PMW-2

Quant Time: Jun 11 03:49:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	137847	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	215605	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214162	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83047	49.121	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.240%
35) Dibromofluoromethane	5.385	113	67136	45.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.080%
50) Toluene-d8	8.647	98	257667	49.469	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.940%
62) 4-Bromofluorobenzene	11.079	95	96847	50.378	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.760%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.470	56	10746	5.384	ug/l	98
39) Methylcyclohexane	7.379	83	16359	7.377	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061024\
Data File : VX041809.D
Acq On : 10 Jun 2024 14:44
Operator : JC/MD
Sample : P2767-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PMW-2

Quant Time: Jun 11 03:49:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

