

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061024\
 Data File : VX041813.D
 Acq On : 10 Jun 2024 16:17
 Operator : JC/MD
 Sample : P2760-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Jun 11 03:50:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	137405	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	215928	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214024	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82392	48.891	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.780%
35) Dibromofluoromethane	5.385	113	67848	45.447	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.900%
50) Toluene-d8	8.646	98	253168	48.533	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.060%
62) 4-Bromofluorobenzene	11.079	95	94884	49.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.560%
Target Compounds						
16) Acetone	2.391	43	4134	2.898	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061024\
Data File : VX041813.D
Acq On : 10 Jun 2024 16:17
Operator : JC/MD
Sample : P2760-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Jun 11 03:50:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

