



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061024\
Data File : VX041796.D
Acq On : 10 Jun 2024 08:47
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Title : SW846 8260
Last Update : Thu May 30 07:36:24 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.2	10118	PASS
75	95	30	60	51.3	27021	PASS
95	95	100	100	100.0	52645	PASS
96	95	5	9	6.7	3524	PASS
173	174	0.00	2	1.3	630	PASS
174	95	50	100	92.4	48624	PASS
175	174	5	9	7.9	3840	PASS
176	174	95	101	96.2	46787	PASS
177	176	5	9	6.4	3012	PASS