

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046587.D
 Acq On : 10 Jun 2025 09:36
 Operator : JC/MD
 Sample : VX0610MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0610MBL01

Quant Time: Jun 11 01:41:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74502	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	143745	50.000	ug/l	-0.01
63) Chlorobenzene-d5	10.049	117	149321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	69510	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	65097	49.113	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.220%
35) Dibromofluoromethane	5.391	113	54195	51.268	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.540%
50) Toluene-d8	8.647	98	191383	54.914	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.820%
62) 4-Bromofluorobenzene	11.079	95	77753	53.922	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	107.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046587.D
Acq On : 10 Jun 2025 09:36
Operator : JC/MD
Sample : VX0610MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0610MBL01

Quant Time: Jun 11 01:41:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

