

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046588.D
 Acq On : 10 Jun 2025 09:57
 Operator : JC/MD
 Sample : VX0610WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0610WBL01

Quant Time: Jun 11 01:42:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	106110	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	211553	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104374	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94499	50.058	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.120%
35) Dibromofluoromethane	5.397	113	77970	50.117	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.240%
50) Toluene-d8	8.647	98	278481	54.294	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.580%
62) 4-Bromofluorobenzene	11.079	95	116202	54.757	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	109.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046588.D
Acq On : 10 Jun 2025 09:57
Operator : JC/MD
Sample : VX0610WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0610WBL01

Quant Time: Jun 11 01:42:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

