

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046591.D
 Acq On : 10 Jun 2025 11:05
 Operator : JC/MD
 Sample : Q2262-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ARS20-0032

Quant Time: Jun 11 01:44:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	108529	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	214443	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	104052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95571	49.497	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.000%
35) Dibromofluoromethane	5.397	113	71641	45.428	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.860%
50) Toluene-d8	8.653	98	279911	53.837	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.680%
62) 4-Bromofluorobenzene	11.079	95	116327	54.077	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.160%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	38100	53.920	ug/l	94
20) Methylene Chloride	2.806	84	4290	2.769	ug/l	87
25) 2-Butanone	4.593	43	6276	5.839	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046591.D
Acq On : 10 Jun 2025 11:05
Operator : JC/MD
Sample : Q2262-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ARS20-0032

Quant Time: Jun 11 01:44:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

