

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046593.D
 Acq On : 10 Jun 2025 11:48
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 11 01:45:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	111878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	217132	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	107755	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	98334	49.404	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.800%
35) Dibromofluoromethane	5.397	113	80392	50.346	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.700%
50) Toluene-d8	8.653	98	289467	54.986	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.980%
62) 4-Bromofluorobenzene	11.079	95	118159	54.248	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.500%
Target Compounds						
16) Acetone	2.392	43	406	5.104	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046593.D
Acq On : 10 Jun 2025 11:48
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 11 01:45:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

