

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046596.D
 Acq On : 10 Jun 2025 12:51
 Operator : JC/MD
 Sample : Q2233-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18B-56-060425DL

Quant Time: Jun 11 01:46:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	113282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	219800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225853	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	112325	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97673	48.463	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.920%
35) Dibromofluoromethane	5.397	113	80813	49.996	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.000%
50) Toluene-d8	8.653	98	290211	54.458	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.920%
62) 4-Bromofluorobenzene	11.079	95	121540	55.123	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.240%
Target Compounds						Qvalue
3) Chloromethane	1.307	50	651	0.455	ug/l	99
4) Vinyl Chloride	1.386	62	4768	3.341	ug/l	97
16) Acetone	2.398	43	1914	6.968	ug/l	98
17) Carbon Disulfide	2.532	76	1157	0.606	ug/l #	84
27) cis-1,2-Dichloroethene	4.501	96	70421	40.192	ug/l	86
44) Trichloroethene	7.135	130	19008	11.509	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046596.D
Acq On : 10 Jun 2025 12:51
Operator : JC/MD
Sample : Q2233-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18B-56-060425DL

Quant Time: Jun 11 01:46:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

