

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046597.D
 Acq On : 10 Jun 2025 13:13
 Operator : JC/MD
 Sample : Q2233-03DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-18B-56-060425-FDDL

Quant Time: Jun 11 01:46:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114089	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221007	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228335	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	113693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99818	49.177	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.360%
35) Dibromofluoromethane	5.403	113	82683	50.873	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.740%
50) Toluene-d8	8.653	98	293804	54.831	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.660%
62) 4-Bromofluorobenzene	11.079	95	123653	55.775	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.560%
Target Compounds						
					Qvalue	
3) Chloromethane	1.307	50	1462	1.014	ug/l	100
4) Vinyl Chloride	1.386	62	6093	4.240	ug/l	93
12) 1,1-Dichloroethene	2.343	96	730	0.539	ug/l	82
16) Acetone	2.392	43	3207	8.544	ug/l #	81
17) Carbon Disulfide	2.532	76	2517	1.074	ug/l	99
27) cis-1,2-Dichloroethene	4.507	96	91409	51.801	ug/l	86
29) Tetrahydrofuran	5.050	42	1290	1.808	ug/l #	71
44) Trichloroethene	7.141	130	12652	7.619	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046597.D
Acq On : 10 Jun 2025 13:13
Operator : JC/MD
Sample : Q2233-03DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-18B-56-060425-FDDL

Quant Time: Jun 11 01:46:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

