

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046598.D
 Acq On : 10 Jun 2025 13:34
 Operator : JC/MD
 Sample : Q2233-04DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-19B-72-060425DL

Quant Time: Jun 11 01:46:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114321	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	223942	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	116361	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	94127	46.279	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.560%
35) Dibromofluoromethane	5.397	113	79249	48.121	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.240%
50) Toluene-d8	8.653	98	297442	54.783	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.560%
62) 4-Bromofluorobenzene	11.079	95	122514	54.537	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	109.080%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.349	101	6291	4.191	ug/l	96
16) Acetone	2.398	43	442	5.138	ug/l	91
27) cis-1,2-Dichloroethene	4.507	96	76787	43.427	ug/l	82
44) Trichloroethene	7.135	130	58002	34.470	ug/l	99
64) Tetrachloroethene	9.275	164	2766	1.791	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046598.D
Acq On : 10 Jun 2025 13:34
Operator : JC/MD
Sample : Q2233-04DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-19B-72-060425DL

Quant Time: Jun 11 01:46:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

