

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046599.D
 Acq On : 10 Jun 2025 13:55
 Operator : JC/MD
 Sample : Q2234-01DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-17B-55-060425DL

Quant Time: Jun 11 01:47:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	110214	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	218656	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	112148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	98868	50.422	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.840%	
35) Dibromofluoromethane	5.397	113	80756	50.222	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.440%	
50) Toluene-d8	8.653	98	288466	54.414	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.820%	
62) 4-Bromofluorobenzene	11.079	95	120075	54.744	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 109.480%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.355	101	976	0.674	ug/l	93
27) cis-1,2-Dichloroethene	4.507	96	27823	16.322	ug/l	86
44) Trichloroethene	7.135	130	61302	37.312	ug/l	98
64) Tetrachloroethene	9.281	164	662	0.442	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046599.D
Acq On : 10 Jun 2025 13:55
Operator : JC/MD
Sample : Q2234-01DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-17B-55-060425DL

Quant Time: Jun 11 01:47:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

