

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046605.D
 Acq On : 10 Jun 2025 16:03
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 11 01:50:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	223910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	116292	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97705	48.106	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.220%
35) Dibromofluoromethane	5.397	113	80574	48.933	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.860%
50) Toluene-d8	8.653	98	293429	54.051	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.100%
62) 4-Bromofluorobenzene	11.079	95	125777	55.998	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	112.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046605.D
Acq On : 10 Jun 2025 16:03
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 11 01:50:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

