

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046623.D
 Acq On : 11 Jun 2025 00:12
 Operator : JC/MD
 Sample : VX0610WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0610WBL02

Quant Time: Jun 11 06:52:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061025W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jun 11 06:20:28 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.916	128	18694	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	99149	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	91863	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.964	65	47443	31.352	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.500%
60) 4-Bromofluorobenzene	11.079	95	46051	29.741	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	99.133%
63) Toluene-d8	8.653	98	123595	30.121	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046623.D
Acq On : 11 Jun 2025 00:12
Operator : JC/MD
Sample : VX0610WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0610WBL02

Quant Time: Jun 11 06:52:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061025W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jun 11 06:20:28 2025
Response via : Initial Calibration

