

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
 Data File : VX046605.D
 Acq On : 10 Jun 2025 16:03
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 11 01:50:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	114161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	223910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231029	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	116292	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97705	48.106	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.220%
35) Dibromofluoromethane	5.397	113	80574	48.933	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.860%
50) Toluene-d8	8.653	98	293429	54.051	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.100%
62) 4-Bromofluorobenzene	11.079	95	125777	55.998	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	112.000%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046605.D
Acq On : 10 Jun 2025 16:03
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 11 01:50:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 114161
Ion Ratio Lower Upper
168 100
99 66.0 54.9 82.3



#33
1,2-Dichloroethane-d4
Concen: 48.106 ug/l
RT: 5.964 min Scan# 800
Delta R.T. -0.006 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03

Tgt Ion: 65 Resp: 97705
Ion Ratio Lower Upper
65 100
67 50.8 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 91
Delta R.T. -0.006 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 223910
Ion Ratio Lower Upper
114 100
63 22.8 0.0 49.2
88 16.9 0.0 33.6



#35
Dibromofluoromethane
Concen: 48.933 ug/l
RT: 5.397 min Scan# 707
Delta R.T. -0.006 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03

Tgt Ion:113 Resp: 80574
Ion Ratio Lower Upper
113 100
111 103.1 83.1 124.7
192 16.9 13.3 19.9









#50
Toluene-d8
Concen: 54.051 ug/l
RT: 8.653 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 293429
Ion Ratio Lower Upper
98 100
100 65.7 53.5 80.3



#62
4-Bromofluorobenzene
Concen: 55.998 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03

Tgt Ion: 95 Resp: 125777
Ion Ratio Lower Upper
95 100
174 69.6 0.0 135.8
176 65.8 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 231029
Ion Ratio Lower Upper
117 100
82 58.0 50.6 76.0
119 32.3 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.006 min
Lab File: VX046605.D
Acq: 10 Jun 2025 16:03

Tgt Ion:152 Resp: 116292
Ion Ratio Lower Upper
152 100
115 64.5 46.9 140.7
150 158.7 0.0 351.0

