

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061119\
 Data File : VX010223.D
 Acq On : 11 Jun 2019 11:35
 Operator : JC/SP
 Sample : K3257-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D3-20190606

Quant Time: Jun 12 08:06:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252271	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	380067	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	340156	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	159551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	111532	45.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.58%	
35) Dibromofluoromethane	5.49	113	116337	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
50) Toluene-d8	8.71	98	442630	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
62) 4-Bromofluorobenzene	11.13	95	150616	45.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.80%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	11729	5.287	ug/l	91
12) 1,1-Dichloroethene	2.37	96	2864	1.264	ug/l #	74
16) Acetone	2.45	43	3363	2.904	ug/l	90
27) cis-1,2-Dichloroethene	4.60	96	2807	1.006	ug/l	76
44) Trichloroethene	7.21	130	575427	199.099	ug/l	87
64) Tetrachloroethene	9.34	164	2847	1.127	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061119\
Data File : VX010223.D
Acq On : 11 Jun 2019 11:35
Operator : JC/SP
Sample : K3257-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D3-20190606

Quant Time: Jun 12 08:06:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

