

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061120\
 Data File : VX016699.D
 Acq On : 11 Jun 2020 13:18
 Operator : JC/SP
 Sample : L2978-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014RW041-200610

Quant Time: Jun 12 07:17:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	207458	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	330564	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	339330	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	172242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	124698	48.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.80%	
35) Dibromofluoromethane	5.47	113	101954	50.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.18%	
50) Toluene-d8	8.70	98	407918	51.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.98%	
62) 4-Bromofluorobenzene	11.12	95	163085	53.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.56%	
Target Compounds						
4) Vinyl Chloride	1.39	62	14611	5.600	ug/l	99
21) trans-1,2-Dichloroethene	3.15	96	1535	0.695	ug/l #	75
27) cis-1,2-Dichloroethene	4.57	96	358778	140.456	ug/l	87
44) Trichloroethene	7.19	130	4214	1.425	ug/l	74
65) Chlorobenzene	10.12	112	41778	6.039	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061120\
Data File : VX016699.D
Acq On : 11 Jun 2020 13:18
Operator : JC/SP
Sample : L2978-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014RW041-200610

Quant Time: Jun 12 07:17:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

