

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061120\
 Data File : VX016700.D
 Acq On : 11 Jun 2020 13:41
 Operator : JC/SP
 Sample : L2978-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014RW040-200610

Quant Time: Jun 12 07:18:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	207177	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	327994	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	337926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	173576	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	123930	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
35) Dibromofluoromethane	5.47	113	100284	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
50) Toluene-d8	8.70	98	404382	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.12	95	164627	54.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.40%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	3312	1.298	ug/l	78
44) Trichloroethene	7.19	130	22335	7.612	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061120\
Data File : VX016700.D
Acq On : 11 Jun 2020 13:41
Operator : JC/SP
Sample : L2978-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014RW040-200610

Quant Time: Jun 12 07:18:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

