

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061120\
 Data File : VX016701.D
 Acq On : 11 Jun 2020 14:03
 Operator : JC/SP
 Sample : L2978-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1-DUP-0569-200610

Quant Time: Jun 12 07:19:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	210707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	329379	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	342374	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	171850	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	125945	48.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.26%	
35) Dibromofluoromethane	5.46	113	101391	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
50) Toluene-d8	8.70	98	405595	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
62) 4-Bromofluorobenzene	11.12	95	163109	53.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.94%	
Target Compounds						
4) Vinyl Chloride	1.39	62	14450	5.453	ug/l	97
21) trans-1,2-Dichloroethene	3.14	96	1569	0.700	ug/l #	72
27) cis-1,2-Dichloroethene	4.57	96	359367	138.517	ug/l	88
44) Trichloroethene	7.19	130	4190	1.422	ug/l	94
65) Chlorobenzene	10.12	112	41559	5.954	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061120\
Data File : VX016701.D
Acq On : 11 Jun 2020 14:03
Operator : JC/SP
Sample : L2978-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1-DUP-0569-200610

Quant Time: Jun 12 07:19:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

