

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061120\
 Data File : VX016702.D
 Acq On : 11 Jun 2020 14:26
 Operator : JC/SP
 Sample : L2977-03DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 E-LGC-2DL

Quant Time: Jun 12 07:21:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	200676	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	323033	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	324785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	177468	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	121771	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	5.46	113	104070	52.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.64%	
50) Toluene-d8	8.70	98	392394	50.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.36%	
62) 4-Bromofluorobenzene	11.12	95	164091	54.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.70%	
Target Compounds						
17) Carbon Disulfide	2.54	76	21993	3.835	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	1462	0.592	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061120\
Data File : VX016702.D
Acq On : 11 Jun 2020 14:26
Operator : JC/SP
Sample : L2977-03DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
E-LGC-2DL

Quant Time: Jun 12 07:21:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

