

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061120\
 Data File : VX016706.D
 Acq On : 11 Jun 2020 15:57
 Operator : JC/SP
 Sample : L2963-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1-DUP-0565-200608

Quant Time: Jun 12 07:36:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	211115	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	329941	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	338168	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	172437	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	125053	47.70	ug/l	0.00
Spiked Amount			Recovery	=	95.40%	
35) Dibromofluoromethane	5.47	113	103209	50.80	ug/l	0.00
Spiked Amount			Recovery	=	101.60%	
50) Toluene-d8	8.70	98	405213	51.24	ug/l	0.00
Spiked Amount			Recovery	=	102.48%	
62) 4-Bromofluorobenzene	11.12	95	162912	53.32	ug/l	0.00
Spiked Amount			Recovery	=	106.64%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.31	50	1548	0.617	ug/l	94
4) Vinyl Chloride	1.39	62	56974	21.458	ug/l	98
12) 1,1-Dichloroethene	2.36	96	1810	0.891	ug/l	93
16) Acetone	2.41	43	3141	2.734	ug/l	94
17) Carbon Disulfide	2.55	76	2441	0.405	ug/l #	85
21) trans-1,2-Dichloroethene	3.15	96	2524	1.123	ug/l	86
27) cis-1,2-Dichloroethene	4.57	96	1332173	512.490	ug/l	88
40) Benzene	6.12	78	3400	0.365	ug/l	100
65) Chlorobenzene	10.12	112	35656	5.172	ug/l	97
91) 1,2-Dichlorobenzene	12.38	146	8472	1.611	ug/l	98
95) Naphthalene	13.82	128	4436	0.370	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061120\
Data File : VX016706.D
Acq On : 11 Jun 2020 15:57
Operator : JC/SP
Sample : L2963-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1-DUP-0565-200608

Quant Time: Jun 12 07:36:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

