

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061120\
 Data File : VX016707.D
 Acq On : 11 Jun 2020 16:20
 Operator : JC/SP
 Sample : L2963-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW028-200608

Quant Time: Jun 12 07:39:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	203661	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	323851	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	334379	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	170802	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	121894	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
35) Dibromofluoromethane	5.47	113	100399	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	8.70	98	400720	51.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.26%	
62) 4-Bromofluorobenzene	11.12	95	160033	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
Target Compounds						
4) Vinyl Chloride	1.39	62	5455	2.130	ug/l	99
16) Acetone	2.42	43	3135	2.829	ug/l	97
17) Carbon Disulfide	2.56	76	1202	0.207	ug/l #	83
27) cis-1,2-Dichloroethene	4.57	96	4210	1.679	ug/l	89
40) Benzene	6.12	78	6771	0.742	ug/l	98
65) Chlorobenzene	10.12	112	949353	139.261	ug/l	99
91) 1,2-Dichlorobenzene	12.38	146	4245	0.815	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061120\
Data File : VX016707.D
Acq On : 11 Jun 2020 16:20
Operator : JC/SP
Sample : L2963-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW028-200608

Quant Time: Jun 12 07:39:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

