

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061120\
 Data File : VX016708.D
 Acq On : 11 Jun 2020 16:42
 Operator : JC/SP
 Sample : L2963-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW015-200608

Quant Time: Jun 12 07:40:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	209252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	328276	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	337989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	173587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	125210	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
35) Dibromofluoromethane	5.47	113	102294	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
50) Toluene-d8	8.70	98	402829	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
62) 4-Bromofluorobenzene	11.12	95	163664	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.39	62	555590	211.114	ug/l	100
27) cis-1,2-Dichloroethene	4.57	96	116831	45.345	ug/l	88
40) Benzene	6.12	78	4703	0.508	ug/l	97
52) Toluene	8.76	92	1349	0.231	ug/l	85
65) Chlorobenzene	10.12	112	69706	10.116	ug/l	99
91) 1,2-Dichlorobenzene	12.38	146	9220	1.741	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061120\
Data File : VX016708.D
Acq On : 11 Jun 2020 16:42
Operator : JC/SP
Sample : L2963-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW015-200608

Quant Time: Jun 12 07:40:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

