

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061120\
 Data File : VX016717.D
 Acq On : 11 Jun 2020 20:06
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 08:04:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.492	0.398	19.1	57	0.00
3 P	Chloromethane	0.595	0.430	27.7#	55	0.00
4 C	Vinyl Chloride	0.629	0.524	16.7#	62	0.00
5 T	Bromomethane	0.306	0.263	14.1	67	0.00
6 T	Chloroethane	0.367	0.339	7.6	68	0.00
7 T	Trichlorofluoromethane	0.826	0.834	-1.0	74	0.00
8 T	Diethyl Ether	0.344	0.328	4.7	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.455	0.454	0.2	74	0.00
10 T	Methyl Iodide	0.490	0.351	28.4#	50#	0.00
11 T	Tert butyl alcohol	0.159	0.149	6.3	69	0.00
12 CM	1,1-Dichloroethene	0.481	0.470	2.3#	73	0.00
13 T	Acrolein	0.068	0.020	70.6#	23#	0.00
14 T	Allyl chloride	0.885	0.741	16.3	62	0.00
15 T	Acrylonitrile	0.326	0.299	8.3	66	0.00
16 T	Acetone	0.272	0.268	1.5	72	0.00
17 T	Carbon Disulfide	1.429	1.188	16.9	62	0.00
18 T	Methyl Acetate	0.698	0.694	0.6	72	0.00
19 T	Methyl tert-butyl Ether	1.728	1.707	1.2	72	0.00
20 T	Methylene Chloride	0.563	0.529	6.0	71	0.00
21 T	trans-1,2-Dichloroethene	0.532	0.499	6.2	70	0.00
22 T	Diisopropyl ether	1.700	1.501	11.7	64	0.00
23 T	Vinyl Acetate	1.488	1.327	10.8	63	0.00
24 P	1,1-Dichloroethane	0.959	0.904	5.7	69	0.00
25 T	2-Butanone	0.461	0.406	11.9	63	0.00
26 T	2,2-Dichloropropane	0.785	0.758	3.4	73	0.00
27 T	cis-1,2-Dichloroethene	0.616	0.587	4.7	71	0.00
28 T	Bromochloromethane	0.446	0.394	11.7	69	0.00
29 T	Tetrahydrofuran	0.298	0.259	13.1	61	0.00
30 C	Chloroform	0.967	0.969	-0.2#	74	0.00
31 T	Cyclohexane	0.863	0.687	20.4	59	0.00
32 T	1,1,1-Trichloroethane	0.875	0.886	-1.3	74	0.00
33 S	1,2-Dichloroethane-d4	0.621	0.609	1.9	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.308	0.329	-6.8	80	0.00
36 T	1,1-Dichloropropene	0.473	0.439	7.2	67	0.00
37 T	Ethyl Acetate	0.568	0.492	13.4	62	0.00
38 T	Carbon Tetrachloride	0.497	0.516	-3.8	74	0.00
39 T	Methylcyclohexane	0.514	0.475	7.6	66	0.00
40 TM	Benzene	1.410	1.341	4.9	68	0.00
41 T	Methacrylonitrile	0.312	0.272	12.8	64	0.00
42 TM	1,2-Dichloroethane	0.501	0.498	0.6	71	0.00
43 T	Isopropyl Acetate	0.887	0.789	11.0	64	0.00
44 TM	Trichloroethene	0.447	0.394	11.9	65	0.00
45 C	1,2-Dichloropropane	0.359	0.345	3.9#	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX061120\
 Data File : VX016717.D
 Acq On : 11 Jun 2020 20:06
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 12 08:04:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.241	0.245	-1.7	73	0.00
47 T	Bromodichloromethane	0.499	0.513	-2.8	73	0.00
48 T	Methyl methacrylate	0.424	0.382	9.9	63	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	64	0.00
50 S	Toluene-d8	1.198	1.213	-1.3	74	0.00
51 T	4-Methyl-2-Pentanone	0.552	0.514	6.9	65	0.00
52 CM	Toluene	0.889	0.879	1.1#	70	0.00
53 T	t-1,3-Dichloropropene	0.568	0.574	-1.1	71	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.594	1.2	70	0.00
55 T	1,1,2-Trichloroethane	0.353	0.369	-4.5	75	0.00
56 T	Ethyl methacrylate	0.576	0.573	0.5	69	0.00
57 T	1,3-Dichloropropane	0.607	0.605	0.3	71	0.00
58 T	2-Chloroethyl Vinyl ether	0.300	0.290	3.3	69	0.00
59 T	2-Hexanone	0.429	0.398	7.2	64	0.00
60 T	Dibromochloromethane	0.398	0.436	-9.5	77	0.00
61 T	1,2-Dibromoethane	0.379	0.395	-4.2	74	0.00
62 S	4-Bromofluorobenzene	0.463	0.480	-3.7	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.559	0.388	30.6#	52	0.00
65 PM	Chlorobenzene	1.019	1.018	0.1	73	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.415	-6.7	79	0.00
67 C	Ethyl Benzene	1.815	1.762	2.9#	70	0.00
68 T	m/p-Xylenes	0.673	0.673	0.0	72	0.00
69 T	o-Xylene	0.652	0.641	1.7	71	0.00
70 T	Styrene	1.120	1.152	-2.9	73	0.00
71 P	Bromoform	0.329	0.369	-12.2	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.464	3.287	5.1	72	0.00
74 T	N-amyl acetate	1.594	1.337	16.1	63	0.00
75 P	1,1,2,2-Tetrachloroethane	0.868	1.051	-21.1	91	0.00
76 T	1,2,3-Trichloropropane	1.099	1.034	5.9	72	0.00
77 T	Bromobenzene	0.909	0.887	2.4	77	0.00
78 T	n-propylbenzene	3.870	3.616	6.6	71	0.00
79 T	2-Chlorotoluene	2.363	2.236	5.4	73	0.00
80 T	1,3,5-Trimethylbenzene	2.893	2.839	1.9	74	0.00
81 T	trans-1,4-Dichloro-2-butene	0.447	0.396	11.4	69	0.00
82 T	4-Chlorotoluene	2.790	2.679	4.0	73	0.00
83 T	tert-Butylbenzene	2.472	2.689	-8.8	82	0.00
84 T	1,2,4-Trimethylbenzene	2.925	2.862	2.2	74	0.00
85 T	sec-Butylbenzene	3.101	3.046	1.8	75	0.00
86 T	p-Isopropyltoluene	2.905	2.916	-0.4	75	0.00
87 T	1,3-Dichlorobenzene	1.574	1.547	1.7	76	0.00
88 T	1,4-Dichlorobenzene	1.610	1.594	1.0	77	0.00
89 T	n-Butylbenzene	2.419	2.371	2.0	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061120\
Data File : VX016717.D
Acq On : 11 Jun 2020 20:06
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 12 08:04:45 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.502	0.511	-1.8	77	0.00
91 T	1,2-Dichlorobenzene	1.525	1.495	2.0	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.281	0.274	2.5	75	0.00
93 T	1,2,4-Trichlorobenzene	1.013	1.078	-6.4	81	0.00
94 T	Hexachlorobutadiene	0.398	0.439	-10.3	88	0.00
95 T	Naphthalene	3.472	3.616	-4.1	78	0.00
96 T	1,2,3-Trichlorobenzene	0.992	1.065	-7.4	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6