

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
 Data File : VX022661.D
 Acq On : 11 Jun 2021 12:47
 Operator : JC/MD
 Sample : M2684-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PILOT-PLANT-PROCESS-BLEED

Quant Time: Jun 11 13:27:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	59964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	119104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102951	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38517	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49284	45.700	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.400%		
35) Dibromofluoromethane	5.397	113	36060	47.453	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.900%		
50) Toluene-d8	8.653	98	144268	48.160	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.320%		
62) 4-Bromofluorobenzene	11.085	95	48972	43.225	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	86.460%		
Target Compounds						Qvalue
11) Tert butyl alcohol	2.953	59	9871	39.532	ug/l #	75
16) Acetone	2.386	43	1380	2.858	ug/l	96
52) Toluene	8.720	92	3497	1.591	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
Data File : VX022661.D
Acq On : 11 Jun 2021 12:47
Operator : JC/MD
Sample : M2684-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PILOT-PLANT-PROCESS-BLEED

Quant Time: Jun 11 13:27:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

