

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
 Data File : VX022667.D
 Acq On : 11 Jun 2021 15:09
 Operator : JC/MD
 Sample : M2688-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB06092021

Quant Time: Jun 12 02:28:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	58733	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	118258	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	105612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	40568	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51297	48.563	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.120%		
35) Dibromofluoromethane	5.397	113	35957	47.656	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.320%		
50) Toluene-d8	8.653	98	144566	48.604	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.200%		
62) 4-Bromofluorobenzene	11.085	95	50335	44.746	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.500%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	5262	11.125	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
Data File : VX022667.D
Acq On : 11 Jun 2021 15:09
Operator : JC/MD
Sample : M2688-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB06092021

Quant Time: Jun 12 02:28:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

