

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
 Data File : VX022672.D
 Acq On : 11 Jun 2021 17:06
 Operator : JC/MD
 Sample : PB136986TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136986TB

Quant Time: Jun 12 02:30:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61991	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	125493	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	116809	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45209	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52113	46.743	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.480%
35) Dibromofluoromethane	5.391	113	36194	45.204	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.400%
50) Toluene-d8	8.653	98	157958	50.045	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.100%
62) 4-Bromofluorobenzene	11.085	95	55716	46.674	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.340%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	6555	13.130	ug/l	# 82
20) Methylene Chloride	2.794	84	2621	2.455	ug/l	# 95
43) Isopropyl Acetate	6.348	43	11239	4.478	ug/l	# 91
52) Toluene	8.726	92	3599	1.554	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	3871	1.234	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
Data File : VX022672.D
Acq On : 11 Jun 2021 17:06
Operator : JC/MD
Sample : PB136986TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136986TB

Quant Time: Jun 12 02:30:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

