

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
 Data File : VX022673.D
 Acq On : 11 Jun 2021 17:30
 Operator : JC/MD
 Sample : M2703-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 210609007-03-TRIP-BLANK

Quant Time: Jun 12 02:30:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60402	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	122955	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	108496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52019	47.886	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.780%
35) Dibromofluoromethane	5.397	113	35692	45.498	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.000%
50) Toluene-d8	8.653	98	149089	48.210	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.420%
62) 4-Bromofluorobenzene	11.085	95	52191	44.624	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.240%
Target Compounds						
30) Chloroform	5.105	83	2697	1.662	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
Data File : VX022673.D
Acq On : 11 Jun 2021 17:30
Operator : JC/MD
Sample : M2703-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
210609007-03-TRIP-BLANK

Quant Time: Jun 12 02:30:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

