

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
 Data File : VX022680.D
 Acq On : 11 Jun 2021 20:15
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 12 02:33:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64928	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	129985	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48437	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	54864	46.984	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.960%
35) Dibromofluoromethane	5.397	113	39158	47.216	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.440%
50) Toluene-d8	8.653	98	161096	49.276	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.560%
62) 4-Bromofluorobenzene	11.085	95	58395	47.228	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.460%
Target Compounds						
95) Naphthalene	13.780	128	1018	1.715	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
Data File : VX022680.D
Acq On : 11 Jun 2021 20:15
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 12 02:33:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

