

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
 Data File : VX022662.D
 Acq On : 11 Jun 2021 13:11
 Operator : JC/MD
 Sample : M2690-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 12 02:27:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60768	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	122280	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109847	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52823	48.333	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.660%
35) Dibromofluoromethane	5.397	113	36541	46.837	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	8.653	98	152994	49.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.500%
62) 4-Bromofluorobenzene	11.085	95	52212	44.888	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.780%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061121\
Data File : VX022662.D
Acq On : 11 Jun 2021 13:11
Operator : JC/MD
Sample : M2690-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 12 02:27:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:168 Resp: 60768
Ion Ratio Lower Upper
168 100
99 73.9 45.8 68.6#



#33
1,2-Dichloroethane-d4
Concen: 48.333 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11

Tgt Ion: 65 Resp: 52823
Ion Ratio Lower Upper
65 100
67 52.1 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:114 Resp: 122280
Ion Ratio Lower Upper
114 100
63 21.1 0.0 40.8
88 17.5 0.0 32.8



#35
Dibromofluoromethane
Concen: 46.837 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11

Tgt Ion:113 Resp: 36541
Ion Ratio Lower Upper
113 100
111 105.4 81.9 122.9
192 15.5 16.7 25.1#









#50
Toluene-d8
Concen: 49.746 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion: 98 Resp: 152994
Ion Ratio Lower Upper
98 100
100 65.5 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 44.888 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11

Tgt Ion: 95 Resp: 52212
Ion Ratio Lower Upper
95 100
174 61.6 0.0 154.6
176 60.2 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tgt Ion:117 Resp: 109847
Ion Ratio Lower Upper
117 100
82 61.4 47.4 71.2
119 32.5 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX022662.D
Acq: 11 Jun 2021 13:11

Tgt Ion:152 Resp: 42054
Ion Ratio Lower Upper
152 100
115 67.8 41.1 123.3
150 163.1 0.0 349.0

