

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046637.D
 Acq On : 11 Jun 2025 15:14
 Operator : JC/MD
 Sample : VX0611WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0611WBL02

Quant Time: Jun 12 02:29:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73470	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.769	114	147517	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	77047	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	64944	49.685	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.380%
35) Dibromofluoromethane	5.391	113	54640	50.367	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.740%
50) Toluene-d8	8.647	98	194299	54.326	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.660%
62) 4-Bromofluorobenzene	11.079	95	81807	55.283	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	110.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046637.D
Acq On : 11 Jun 2025 15:14
Operator : JC/MD
Sample : VX0611WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0611WBL02

Quant Time: Jun 12 02:29:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

