

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046638.D
 Acq On : 11 Jun 2025 15:36
 Operator : JC/MD
 Sample : VX0611MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0611MBL01

Quant Time: Jun 12 02:30:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	111144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222596	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	117551	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	99780	50.461	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.920%
35) Dibromofluoromethane	5.403	113	82721	50.533	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.060%
50) Toluene-d8	8.653	98	293378	54.361	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	108.720%
62) 4-Bromofluorobenzene	11.079	95	126910	56.836	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	113.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046638.D
Acq On : 11 Jun 2025 15:36
Operator : JC/MD
Sample : VX0611MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0611MBL01

Quant Time: Jun 12 02:30:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

