

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046643.D
 Acq On : 11 Jun 2025 17:31
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 12 02:34:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	113796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	224937	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	116079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	96176	47.505	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.000%
35) Dibromofluoromethane	5.403	113	81859	49.486	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.980%
50) Toluene-d8	8.653	98	288717	52.940	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.880%
62) 4-Bromofluorobenzene	11.079	95	126916	56.247	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	112.500%
Target Compounds						
16) Acetone	2.404	43	917	5.727	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046643.D
Acq On : 11 Jun 2025 17:31
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 12 02:34:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

