

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046644.D
 Acq On : 11 Jun 2025 17:53
 Operator : JC/MD
 Sample : Q2249-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-06-6.5-060525

Quant Time: Jun 12 02:34:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.574	168	113972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	227958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	124632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	86239	42.531	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.060%
35) Dibromofluoromethane	5.410	113	71170	42.454	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.900%
50) Toluene-d8	8.653	98	295147	53.402	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.800%
62) 4-Bromofluorobenzene	11.079	95	126952	55.517	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.040%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.349	101	37948	25.356	ug/l	91
27) cis-1,2-Dichloroethene	4.513	96	2501	1.419	ug/l	84
44) Trichloroethene	7.135	130	227940	133.077	ug/l	98
64) Tetrachloroethene	9.275	164	1879	1.185	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046644.D
Acq On : 11 Jun 2025 17:53
Operator : JC/MD
Sample : Q2249-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-06-6.5-060525

Quant Time: Jun 12 02:34:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

