

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046647.D
 Acq On : 11 Jun 2025 18:59
 Operator : JC/MD
 Sample : Q2273-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-4

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Mahesh Dadoda 06/12/2025

Quant Time: Jun 12 02:35:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	122059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	234404	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245538	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	126422	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	90118	41.499	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	83.000%		
35) Dibromofluoromethane	5.403	113	85156	49.400	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.800%		
50) Toluene-d8	8.653	98	309815	54.515	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	109.020%		
62) 4-Bromofluorobenzene	11.079	95	129279	54.980	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	109.960%		
Target Compounds						
16) Acetone	2.398	43	10842	17.075	ug/l	90
18) Methyl Acetate	2.721	43	10156	3.926	ug/l	97
20) Methylene Chloride	2.812	84	3889	2.232	ug/l #	78
37) Ethyl Acetate	4.751	43	5999m	2.496	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046647.D
Acq On : 11 Jun 2025 18:59
Operator : JC/MD
Sample : Q2273-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-4

Quant Time: Jun 12 02:35:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/12/2025
Supervised By : Mahesh Dadoda 06/12/2025

