

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046649.D
 Acq On : 11 Jun 2025 19:43
 Operator : JC/MD
 Sample : Q2274-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-13-MHP-WC

Quant Time: Jun 12 02:36:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Mahesh Dadoda 06/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	117684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	198220	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	126377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	92861	44.352	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 88.700%			
35) Dibromofluoromethane	5.404	113	80740	55.389	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.780%			
50) Toluene-d8	8.653	98	289567	60.253	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 120.500%#			
62) 4-Bromofluorobenzene	11.079	95	126731	63.735	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 127.480%#			
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	10206	16.780	ug/l #	85
18) Methyl Acetate	2.721	43	10369	4.158	ug/l	100
20) Methylene Chloride	2.806	84	3622	2.156	ug/l	92
37) Ethyl Acetate	4.745	43	5960m	2.933	ug/l	
43) Isopropyl Acetate	6.361	43	3710	1.047	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046649.D
Acq On : 11 Jun 2025 19:43
Operator : JC/MD
Sample : Q2274-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-13-MHP-WC

Quant Time: Jun 12 02:36:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/12/2025
Supervised By : Mahesh Dadoda 06/12/2025

