

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046650.D
 Acq On : 11 Jun 2025 20:04
 Operator : JC/MD
 Sample : Q2278-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-2

Quant Time: Jun 12 02:37:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Mahesh Dadoda 06/12/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	80049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	147113	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176561	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	90971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	62562	43.929	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	87.860%		
35) Dibromofluoromethane	5.403	113	58330	53.916	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	107.840%		
50) Toluene-d8	8.653	98	215519	60.424	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	120.840%#		
62) 4-Bromofluorobenzene	11.079	95	90785	61.519	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	123.040%#		
Target Compounds					Qvalue	
16) Acetone	2.392	43	11190	24.236	ug/l	100
18) Methyl Acetate	2.721	43	9764	5.756	ug/l	98
20) Methylene Chloride	2.806	84	4821	4.220	ug/l #	83
37) Ethyl Acetate	4.745	43	5525m	3.663	ug/l	
43) Isopropyl Acetate	6.367	43	4176	1.588	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046650.D
Acq On : 11 Jun 2025 20:04
Operator : JC/MD
Sample : Q2278-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-2

Quant Time: Jun 12 02:37:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/12/2025
Supervised By :Mahesh Dadoda 06/12/2025

