

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
 Data File : VX046651.D
 Acq On : 11 Jun 2025 20:26
 Operator : JC/MD
 Sample : Q2280-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-210

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/12/2025
 Supervised By : Mahesh Dadoda 06/12/2025

Quant Time: Jun 12 02:37:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 06 16:56:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	121242	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	226103	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	129063	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	88988	41.255	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	82.520%		
35) Dibromofluoromethane	5.403	113	82446	49.584	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.160%		
50) Toluene-d8	8.653	98	307056	56.013	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	112.020%		
62) 4-Bromofluorobenzene	11.079	95	129178	56.954	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	113.900%		
Target Compounds						
16) Acetone	2.398	43	13551	20.299	ug/l	95
18) Methyl Acetate	2.721	43	9963	3.878	ug/l	99
20) Methylene Chloride	2.806	84	4021	2.324	ug/l	89
37) Ethyl Acetate	4.757	43	4500m	1.941	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046651.D
Acq On : 11 Jun 2025 20:26
Operator : JC/MD
Sample : Q2280-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-210

Quant Time: Jun 12 02:37:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 06 16:56:12 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 06/12/2025
Supervised By : Mahesh Dadoda 06/12/2025

