

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061219\
 Data File : VX010243.D
 Acq On : 12 Jun 2019 12:17
 Operator : JC/SP
 Sample : K3309-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1-TB12-190610

Quant Time: Jun 13 01:53:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	330613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	500493	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	451907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	209804	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	149763	46.41	ug/l	0.00
Spiked Amount			Recovery	=	92.82%	
35) Dibromofluoromethane	5.50	113	153944	49.87	ug/l	0.00
Spiked Amount			Recovery	=	99.74%	
50) Toluene-d8	8.71	98	584940	50.56	ug/l	0.00
Spiked Amount			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	199315	46.13	ug/l	0.00
Spiked Amount			Recovery	=	92.26%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061219\
Data File : VX010243.D
Acq On : 12 Jun 2019 12:17
Operator : JC/SP
Sample : K3309-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1-TB12-190610

Quant Time: Jun 13 01:53:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

