

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061219\
 Data File : VX010248.D
 Acq On : 12 Jun 2019 14:14
 Operator : JC/SP
 Sample : K3309-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF017MW003-190610

Quant Time: Jun 13 02:07:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	316123	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	474367	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	429039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199648	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	141873	45.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.96%	
35) Dibromofluoromethane	5.50	113	146571	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
50) Toluene-d8	8.71	98	558500	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	
62) 4-Bromofluorobenzene	11.13	95	187531	45.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.58%	
Target Compounds						
4) Vinyl Chloride	1.41	62	1471	0.542	ug/l	96
21) trans-1,2-Dichloroethene	3.17	96	13724	4.367	ug/l #	76
27) cis-1,2-Dichloroethene	4.60	96	77388	22.138	ug/l	68
44) Trichloroethene	7.21	130	46716	12.951	ug/l	88
64) Tetrachloroethene	9.34	164	47713	14.973	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061219\
Data File : VX010248.D
Acq On : 12 Jun 2019 14:14
Operator : JC/SP
Sample : K3309-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF017MW003-190610

Quant Time: Jun 13 02:07:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

