

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061219\
 Data File : VX010254.D
 Acq On : 12 Jun 2019 16:35
 Operator : JC/SP
 Sample : K3309-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS032MW001-190611

Quant Time: Jun 13 02:30:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	306613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	462431	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	418824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193918	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	140466	46.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.86%	
35) Dibromofluoromethane	5.50	113	141179	49.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.00%	
50) Toluene-d8	8.71	98	542787	50.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.54%	
62) 4-Bromofluorobenzene	11.13	95	184002	46.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061219\
Data File : VX010254.D
Acq On : 12 Jun 2019 16:35
Operator : JC/SP
Sample : K3309-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS032MW001-190611

Quant Time: Jun 13 02:30:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

