

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061219\
 Data File : VX010262.D
 Acq On : 12 Jun 2019 19:41
 Operator : JC/SP
 Sample : K3315-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE104D3-20190607

Quant Time: Jun 13 03:42:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	301950	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	461659	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	419938	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	195646	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	139165	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
35) Dibromofluoromethane	5.50	113	140556	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	8.71	98	542064	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
62) 4-Bromofluorobenzene	11.13	95	185269	46.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.96%	
Target Compounds						
4) Vinyl Chloride	1.41	62	18521	7.142	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061219\
Data File : VX010262.D
Acq On : 12 Jun 2019 19:41
Operator : JC/SP
Sample : K3315-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE104D3-20190607

Quant Time: Jun 13 03:42:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

