

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061219\
 Data File : VX010239.D
 Acq On : 12 Jun 2019 09:46
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 13 01:42:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	-0.01
2 T	Dichlorodifluoromethane	0.460	0.471	-2.4	95	0.00
3 P	Chloromethane	0.457	0.416	9.0	92	0.00
4 C	Vinyl Chloride	0.429	0.467	-8.9#	110	0.00
5 T	Bromomethane	0.162	0.186	-14.8	117	0.00
6 T	Chloroethane	0.217	0.271	-24.9#	114	0.00
7 T	Trichlorofluoromethane	0.638	0.723	-13.3	116	0.00
8 T	Diethyl Ether	0.220	0.255	-15.9	123	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.440	0.431	2.0	103	0.00
10 T	Methyl Iodide	0.681	0.614	9.8	89	0.00
11 T	Tert butyl alcohol	0.120	0.113	5.8	96	0.00
12 CM	1,1-Dichloroethene	0.449	0.413	8.0#	95	0.00
13 T	Acrolein	0.022	0.023	-4.5	99	0.00
14 T	Allyl chloride	0.678	0.612	9.7	94	0.00
15 T	Acrylonitrile	0.244	0.232	4.9	97	0.00
16 T	Acetone	0.230	0.273	-18.7	124	0.00
17 T	Carbon Disulfide	1.298	1.102	15.1	91	0.00
18 T	Methyl Acetate	0.505	0.479	5.1	101	0.00
19 T	Methyl tert-butyl Ether	1.443	1.348	6.6	97	0.00
20 T	Methylene Chloride	0.504	0.466	7.5	98	0.00
21 T	trans-1,2-Dichloroethene	0.497	0.442	11.1	95	0.00
22 T	Diisopropyl ether	1.296	1.263	2.5	99	0.00
23 T	Vinyl Acetate	1.156	1.117	3.4	97	0.00
24 P	1,1-Dichloroethane	0.787	0.732	7.0	98	0.00
25 T	2-Butanone	0.342	0.365	-6.7	109	0.00
26 T	2,2-Dichloropropane	0.718	0.671	6.5	97	0.00
27 T	cis-1,2-Dichloroethene	0.553	0.526	4.9	99	0.00
28 T	Bromochloromethane	0.353	0.327	7.4	97	0.00
29 T	Tetrahydrofuran	0.215	0.204	5.1	98	0.00
30 C	Chloroform	0.817	0.777	4.9#	98	0.00
31 T	Cyclohexane	0.710	0.683	3.8	98	0.00
32 T	1,1,1-Trichloroethane	0.768	0.710	7.6	98	0.00
33 S	1,2-Dichloroethane-d4	0.488	0.450	7.8	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.308	0.309	-0.3	99	0.00
36 T	1,1-Dichloropropene	0.408	0.395	3.2	98	0.00
37 T	Ethyl Acetate	0.424	0.409	3.5	97	0.00
38 T	Carbon Tetrachloride	0.462	0.447	3.2	98	0.00
39 T	Methylcyclohexane	0.555	0.545	1.8	100	0.00
40 TM	Benzene	1.261	1.241	1.6	98	0.00
41 T	Methacrylonitrile	0.230	0.222	3.5	96	0.00
42 TM	1,2-Dichloroethane	0.403	0.381	5.5	95	0.00
43 T	Isopropyl Acetate	0.710	0.687	3.2	97	0.00
44 TM	Trichloroethene	0.380	0.375	1.3	101	0.00
45 C	1,2-Dichloropropane	0.316	0.310	1.9#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061219\
 Data File : VX010239.D
 Acq On : 12 Jun 2019 09:46
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 13 01:42:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.237	0.226	4.6	99	0.00
47 T	Bromodichloromethane	0.442	0.434	1.8	98	0.00
48 T	Methyl methacrylate	0.332	0.324	2.4	95	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	99	0.00
50 S	Toluene-d8	1.156	1.184	-2.4	100	0.00
51 T	4-Methyl-2-Pentanone	0.424	0.438	-3.3	100	0.00
52 CM	Toluene	0.829	0.827	0.2#	101	0.00
53 T	t-1,3-Dichloropropene	0.511	0.491	3.9	95	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.530	3.3	97	0.00
55 T	1,1,2-Trichloroethane	0.340	0.351	-3.2	102	0.00
56 T	Ethyl methacrylate	0.536	0.539	-0.6	98	0.00
57 T	1,3-Dichloropropane	0.535	0.524	2.1	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.253	0.237	6.3	90	0.00
59 T	2-Hexanone	0.338	0.362	-7.1	104	0.00
60 T	Dibromochloromethane	0.402	0.409	-1.7	101	0.00
61 T	1,2-Dibromoethane	0.378	0.374	1.1	99	0.00
62 S	4-Bromofluorobenzene	0.432	0.443	-2.5	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.371	0.374	-0.8	105	0.00
65 PM	Chlorobenzene	1.032	1.028	0.4	101	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.396	0.3	101	0.00
67 C	Ethyl Benzene	1.757	1.756	0.1#	101	0.00
68 T	m/p-Xylenes	0.680	0.688	-1.2	102	0.00
69 T	o-Xylene	0.676	0.674	0.3	102	0.00
70 T	Styrene	1.145	1.168	-2.0	102	0.00
71 P	Bromoform	0.365	0.372	-1.9	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.490	3.418	2.1	104	0.00
74 T	N-amyl acetate	1.380	1.318	4.5	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.195	1.136	4.9	102	0.00
76 T	1,2,3-Trichloropropane	1.000	0.958	4.2	102	0.00
77 T	Bromobenzene	0.960	0.939	2.2	104	0.00
78 T	n-propylbenzene	3.886	3.856	0.8	104	0.00
79 T	2-Chlorotoluene	2.334	2.242	3.9	102	0.00
80 T	1,3,5-Trimethylbenzene	2.940	2.872	2.3	104	0.00
81 T	trans-1,4-Dichloro-2-butene	0.453	0.424	6.4	98	0.00
82 T	4-Chlorotoluene	2.651	2.570	3.1	101	0.00
83 T	tert-Butylbenzene	2.975	2.937	1.3	105	0.00
84 T	1,2,4-Trimethylbenzene	2.969	2.913	1.9	104	0.00
85 T	sec-Butylbenzene	3.467	3.460	0.2	106	0.00
86 T	p-Isopropyltoluene	3.169	3.207	-1.2	107	0.00
87 T	1,3-Dichlorobenzene	1.691	1.690	0.1	106	0.00
88 T	1,4-Dichlorobenzene	1.704	1.690	0.8	108	0.00
89 T	n-Butylbenzene	2.728	2.792	-2.3	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061219\
Data File : VX010239.D
Acq On : 12 Jun 2019 09:46
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 13 01:42:13 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.631	0.605	4.1	102	0.00
91 T	1,2-Dichlorobenzene	1.644	1.646	-0.1	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.281	0.241	14.2	94	0.00
93 T	1,2,4-Trichlorobenzene	1.203	1.201	0.2	109	0.00
94 T	Hexachlorobutadiene	0.535	0.587	-9.7	120	0.00
95 T	Naphthalene	3.787	3.689	2.6	104	0.00
96 T	1,2,3-Trichlorobenzene	1.190	1.194	-0.3	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6