

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061219\
 Data File : VX010240.D
 Acq On : 12 Jun 2019 10:27
 Operator : JC/SP
 Sample : VX0612WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0612WBL01

Quant Time: Jun 13 01:46:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	332123	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	507415	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	450083	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	211583	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	157438	48.56	ug/l	0.00
Spiked Amount			Recovery	=	97.12%	
35) Dibromofluoromethane	5.49	113	164384	52.52	ug/l	0.00
Spiked Amount			Recovery	=	105.04%	
50) Toluene-d8	8.71	98	624402	53.23	ug/l	0.00
Spiked Amount			Recovery	=	106.46%	
62) 4-Bromofluorobenzene	11.13	95	214279	48.91	ug/l	0.00
Spiked Amount			Recovery	=	97.82%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061219\
Data File : VX010240.D
Acq On : 12 Jun 2019 10:27
Operator : JC/SP
Sample : VX0612WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0612WBL01

Quant Time: Jun 13 01:46:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX010240.D

Acq: 12 Jun 2019 10:27

Instrument :

MSVOA_X

ClientSampled :


VX0612WBL01

Tot Ion:168 Resp: 332123

Ion Ratio Lower Upper

168 100

99 46.6 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



5.66

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 48.564 ug/l

RT: 6.05 min Scan# 804

Delta R.T. -0.01 min

Lab File: VX010240.D


Acq: 12 Jun 2019 10:27

Tgt Ion: 65 Resp: 157438

Ion Ratio Lower Upper

65 100

67 55.3 0.0 108.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. -0.01 min

Lab File: VX010240.D

Acq: 12 Jun 2019 10:27

Instrument :

MSVOA_X

ClientSampleId :

VX0612WBL01


Tot Ion:114 Resp: 507415

Ion Ratio Lower Upper

114 100

63 16.4 0.0 47.4




88 14.3 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 52.521 ug/l

RT: 5.49 min Scan# 712

Delta R.T. -0.01 min

Lab File: VX010240.D

Acq: 12 Jun 2019 10:27


Tgt Ion:113 Resp: 164384

Ion Ratio Lower Upper

113 100

111 103.1 82.2 123.2



192 23.0 17.1 25.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 53.230 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX010240.D
Acq: 12 Jun 2019 10:27




Instrument :
MSVOA_X
ClientSampleId :
VX0612WBL01


Tot Ion: 98 Resp: 624402
Ion Ratio Lower Upper
98 100
100 64.4 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 48.914 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. 0.00 min
Lab File: VX010240.D
Acq: 12 Jun 2019 10:27

Tgt Ion: 95 Resp: 214279
Ion Ratio Lower Upper
95 100
174 93.6 0.0 160.4
176 90.7 0.0 154.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX010240.D
Acq: 12 Jun 2019 10:27

Instrument :
MSVOA_X
ClientSampled :
VX0612WBL01

Tot Ion:117 Resp: 450083
Ion Ratio Lower Upper
117 100
82 49.7 49.2 73.8
119 32.4 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX010240.D
Acq: 12 Jun 2019 10:27

Tgt Ion:152 Resp: 211583
Ion Ratio Lower Upper
152 100
115 54.0 41.4 124.2
150 156.4 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

