

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061219\
 Data File : VX010246.D
 Acq On : 12 Jun 2019 13:28
 Operator : JC/SP
 Sample : K3315-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02-20190610

Quant Time: Jun 13 02:03:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	326787	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	493191	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	445254	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	211931	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	143968	45.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.26%	
35) Dibromofluoromethane	5.50	113	150638	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
50) Toluene-d8	8.71	98	574485	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
62) 4-Bromofluorobenzene	11.13	95	198010	46.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.00%	
Target Compounds						
16) Acetone	2.45	43	10800	7.199	ug/l	90
52) Toluene	8.79	92	16173	1.977	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061219\
Data File : VX010246.D
Acq On : 12 Jun 2019 13:28
Operator : JC/SP
Sample : K3315-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02-20190610

Quant Time: Jun 13 02:03:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX010246.D

Acq: 12 Jun 2019 13:28

Instrument :

MSVOA_X

ClientSampleId :


FB02-20190610

Tot Ion:168 Resp: 326787

Ion Ratio Lower Upper

168 100

99 46.1 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



5.67

100000

50000

0

Time--> 5.50 5.60 5.70 5.80



#16

Acetone

Concen: 7.199 ug/l

RT: 2.45 min Scan# 212

Delta R.T. -0.00 min

Lab File: VX010246.D


Acq: 12 Jun 2019 13:28

Tgt Ion: 43 Resp: 10800

Ion Ratio Lower Upper

43 100

58 34.5 23.2 34.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0

2.45

8000


6000


4000

2000

0

Time--> 2.40 2.45 2.50





#33

1,2-Dichloroethane-d4

Concen: 45.134 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.01 min

Lab File: VX010246.D

Acq: 12 Jun 2019 13:28

Instrument :

MSVOA_X

ClientSampleId :





FB02-20190610

Tot Ion: 65 Resp: 143968

Ion Ratio Lower Upper

65 100

67 56.0 0.0 108.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX010246.D

Acq: 12 Jun 2019 13:28




Tgt Ion: 114 Resp: 493191


Ion Ratio Lower Upper

114 100

63 16.8 0.0 47.4

88 14.5 0.0 33.0





#35

Dibromofluoromethane

Concen: 49.517 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX010246.D

Acq: 12 Jun 2019 13:28

Instrument :

MSVOA_X

ClientSampleId :

FB02-20190610


Tot Ion:113 Resp: 150638

Ion Ratio Lower Upper

113 100

111 102.7 82.2 123.2




192 23.6 17.1 25.7



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#50

Toluene-d8

Concen: 50.387 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX010246.D


Acq: 12 Jun 2019 13:28

Tgt Ion: 98 Resp: 574485

Ion Ratio Lower Upper



98 100


100 64.8 51.5 77.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V





#52

Toluene

Concen: 1.977 ug/l

RT: 8.79 min Scan# 1252

Delta R.T. 0.01 min

Lab File: VX010246.D

Acq: 12 Jun 2019 13:28

Instrument :

MSVOA_X

ClientSampleId :





FB02-20190610

Tot Ion: 92 Resp: 16173

Ion Ratio Lower Upper

92 100

91 175.0 139.9 209.9



#62

4-Bromofluorobenzene

Concen: 46.504 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX010246.D

Acq: 12 Jun 2019 13:28




Tgt Ion: 95 Resp: 198010


Ion Ratio Lower Upper

95 100

174 94.2 0.0 160.4

176 92.0 0.0 154.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX010246.D
Acq: 12 Jun 2019 13:28

Instrument :
MSVOA_X
ClientSampleId :
FB02-20190610

Tot Ion:117 Resp: 445254
Ion Ratio Lower Upper
117 100
82 49.6 49.2 73.8
119 31.4 25.2 37.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX010246.D
Acq: 12 Jun 2019 13:28

Tgt Ion:152 Resp: 211931
Ion Ratio Lower Upper
152 100
115 53.9 41.4 124.2
150 155.8 0.0 345.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

