


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061219\
Data File : VX010238.D
Acq On : 12 Jun 2019 09:12
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Title : SW846 8260
Last Update : Sat Jun 08 02:12:13 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1627

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.9	11793	PASS
75	95	30	60	45.7	33856	PASS
95	95	100	100	100.0	74104	PASS
96	95	5	9	6.7	4958	PASS
173	174	0.00	2	1.1	786	PASS
174	95	50	100	92.6	68592	PASS
175	174	5	9	7.6	5228	PASS
176	174	95	101	96.5	66170	PASS
177	176	5	9	6.5	4290	PASS