

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061220\  
 Data File : VX016743.D  
 Acq On : 12 Jun 2020 20:50  
 Operator : JC/SP  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :

Quant Time: Jun 13 04:36:37 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X052720W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 27 16:31:05 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.63  | 168  | 228497   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.83  | 114  | 342987   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.10 | 117  | 328203   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.06 | 152  | 177131   | 50.00 | ug/l  | 0.00     |

## System Monitoring Compounds

|                           |        |     |          |       |         |      |
|---------------------------|--------|-----|----------|-------|---------|------|
| 33) 1,2-Dichloroethane-d4 | 6.03   | 65  | 135121   | 47.62 | ug/l    | 0.00 |
| Spiked Amount             | 50.000 |     | Recovery | =     | 95.24%  |      |
| 35) Dibromofluoromethane  | 5.47   | 113 | 112178   | 53.12 | ug/l    | 0.00 |
| Spiked Amount             | 50.000 |     | Recovery | =     | 106.24% |      |
| 50) Toluene-d8            | 8.70   | 98  | 412560   | 50.19 | ug/l    | 0.00 |
| Spiked Amount             | 50.000 |     | Recovery | =     | 100.38% |      |
| 62) 4-Bromofluorobenzene  | 11.12  | 95  | 164257   | 51.71 | ug/l    | 0.00 |
| Spiked Amount             | 50.000 |     | Recovery | =     | 103.42% |      |

## Target Compounds

|                               |      |     |         |         |        | Qvalue |
|-------------------------------|------|-----|---------|---------|--------|--------|
| 2) Dichlorodifluoromethane    | 1.18 | 85  | 96914   | 43.099  | ug/l   | 97     |
| 3) Chloromethane              | 1.31 | 50  | 101834  | 37.475  | ug/l   | 100    |
| 4) Vinyl Chloride             | 1.39 | 62  | 122966  | 42.789  | ug/l   | 100    |
| 5) Bromomethane               | 1.62 | 94  | 66974   | 47.932  | ug/l   | 99     |
| 6) Chloroethane               | 1.70 | 64  | 79901   | 47.664  | ug/l   | 98     |
| 7) Trichlorofluoromethane     | 1.91 | 101 | 207430  | 54.962  | ug/l   | 97     |
| 8) Diethyl Ether              | 2.17 | 74  | 78107   | 49.633  | ug/l   | 93     |
| 9) 1,1,2-Trichlorotrifluoroet | 2.36 | 101 | 111586  | 53.722  | ug/l   | 96     |
| 10) Methyl Iodide             | 2.49 | 142 | 112477  | 45.610  | ug/l   | 99     |
| 11) Tert butyl alcohol        | 3.03 | 59  | 169489  | 232.924 | ug/l   | 100    |
| 12) 1,1-Dichloroethene        | 2.35 | 96  | 110845  | 50.415  | ug/l   | 98     |
| 13) Acrolein                  | 2.27 | 56  | 80144   | 258.916 | ug/l   | 98     |
| 14) Allyl chloride            | 2.70 | 41  | 174266  | 43.099  | ug/l   | 96     |
| 15) Acrylonitrile             | 3.12 | 53  | 356790  | 239.366 | ug/l   | 99     |
| 16) Acetone                   | 2.42 | 43  | 300352  | 241.589 | ug/l   | 99     |
| 17) Carbon Disulfide          | 2.54 | 76  | 274546  | 42.049  | ug/l   | 98     |
| 18) Methyl Acetate            | 2.75 | 43  | 176385  | 55.316  | ug/l   | 94     |
| 19) Methyl tert-butyl Ether   | 3.17 | 73  | 405327  | 51.340  | ug/l   | 96     |
| 20) Methylene Chloride        | 2.83 | 84  | 124204  | 48.287  | ug/l   | 97     |
| 21) trans-1,2-Dichloroethene  | 3.14 | 96  | 118111  | 48.567  | ug/l   | 91     |
| 22) Diisopropyl ether         | 3.82 | 45  | 349193  | 44.955  | ug/l   | 96     |
| 23) Vinyl Acetate             | 3.78 | 43  | 1502925 | 221.053 | ug/l   | 97     |
| 24) 1,1-Dichloroethane        | 3.67 | 63  | 210015  | 47.904  | ug/l   | 100    |
| 25) 2-Butanone                | 4.64 | 43  | 479682  | 227.703 | ug/l   | 96     |
| 26) 2,2-Dichloropropane       | 4.54 | 77  | 168392  | 46.939  | ug/l   | 98     |
| 27) cis-1,2-Dichloroethene    | 4.56 | 96  | 137622  | 48.916  | ug/l   | 96     |
| 28) Bromochloromethane        | 4.98 | 49  | 83202   | 40.839  | ug/l # | 82     |
| 29) Tetrahydrofuran           | 5.09 | 42  | 308660  | 226.794 | ug/l   | 93     |
| 30) Chloroform                | 5.18 | 83  | 228348  | 51.692  | ug/l   | 98     |
| 31) Cyclohexane               | 5.55 | 56  | 168224  | 42.652  | ug/l   | 97     |
| 32) 1,1,1-Trichloroethane     | 5.46 | 97  | 212042  | 53.020  | ug/l   | 98     |
| 36) 1,1-Dichloropropene       | 5.77 | 75  | 160955  | 49.564  | ug/l   | 97     |
| 37) Ethyl Acetate             | 4.79 | 43  | 174486  | 44.754  | ug/l   | 97     |
| 38) Carbon Tetrachloride      | 5.75 | 117 | 191415  | 56.089  | ug/l   | 97     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061220\  
 Data File : VX016743.D  
 Acq On : 12 Jun 2020 20:50  
 Operator : JC/SP  
 Sample : VSTDCCC050  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :

Quant Time: Jun 13 04:36:37 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X052720W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 27 16:31:05 2020  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc    | Units | Dev(Min) |
|--------------------------------|-------|------|----------|---------|-------|----------|
| 39) Methylcyclohexane          | 7.44  | 83   | 175299   | 49.698  | ug/l  | 99       |
| 40) Benzene                    | 6.11  | 78   | 484622   | 50.112  | ug/l  | 98       |
| 41) Methacrylonitrile          | 5.01  | 41   | 97276    | 45.423  | ug/l  | 95       |
| 42) 1,2-Dichloroethane         | 6.16  | 62   | 178412   | 51.877  | ug/l  | 100      |
| 43) Isopropyl Acetate          | 6.41  | 43   | 282538   | 46.459  | ug/l  | 96       |
| 44) Trichloroethene            | 7.19  | 130  | 148419   | 48.372  | ug/l  | 95       |
| 45) 1,2-Dichloropropane        | 7.49  | 63   | 122279   | 49.668  | ug/l  | 99       |
| 46) Dibromomethane             | 7.64  | 93   | 87458    | 52.889  | ug/l  | 96       |
| 47) Bromodichloromethane       | 7.88  | 83   | 188603   | 55.084  | ug/l  | 100      |
| 48) Methyl methacrylate        | 7.74  | 41   | 138749   | 47.650  | ug/l  | 94       |
| 49) 1,4-Dioxane                | 7.71  | 88   | 63483    | 931.606 | ug/l  | 97       |
| 51) 4-Methyl-2-Pentanone       | 8.62  | 43   | 924833   | 244.232 | ug/l  | 97       |
| 52) Toluene                    | 8.77  | 92   | 318820   | 52.280  | ug/l  | 100      |
| 53) t-1,3-Dichloropropene      | 9.02  | 75   | 206798   | 53.115  | ug/l  | 99       |
| 54) cis-1,3-Dichloropropene    | 8.42  | 75   | 213542   | 51.818  | ug/l  | 98       |
| 55) 1,1,2-Trichloroethane      | 9.20  | 97   | 133729   | 55.153  | ug/l  | 96       |
| 56) Ethyl methacrylate         | 9.16  | 69   | 205845   | 52.127  | ug/l  | 96       |
| 57) 1,3-Dichloropropane        | 9.35  | 76   | 219628   | 52.772  | ug/l  | 100      |
| 58) 2-Chloroethyl Vinyl ether  | 8.29  | 63   | 488990   | 237.239 | ug/l  | 95       |
| 59) 2-Hexanone                 | 9.48  | 43   | 720883   | 244.803 | ug/l  | 97       |
| 60) Dibromochloromethane       | 9.57  | 129  | 160390   | 58.780  | ug/l  | 100      |
| 61) 1,2-Dibromoethane          | 9.65  | 107  | 145363   | 55.985  | ug/l  | 98       |
| 64) Tetrachloroethene          | 9.32  | 164  | 152252   | 41.464  | ug/l  | 96       |
| 65) Chlorobenzene              | 10.12 | 112  | 354446   | 52.972  | ug/l  | 100      |
| 66) 1,1,1,2-Tetrachloroethane  | 10.21 | 131  | 143827   | 56.310  | ug/l  | 98       |
| 67) Ethyl Benzene              | 10.24 | 91   | 613407   | 51.490  | ug/l  | 100      |
| 68) m/p-Xylenes                | 10.34 | 106  | 469251   | 106.237 | ug/l  | 99       |
| 69) o-Xylene                   | 10.68 | 106  | 229136   | 53.565  | ug/l  | 97       |
| 70) Styrene                    | 10.70 | 104  | 409050   | 55.632  | ug/l  | 99       |
| 71) Bromoform                  | 10.84 | 173  | 130412   | 60.414  | ug/l  | 98       |
| 73) Isopropylbenzene           | 11.00 | 105  | 614309   | 50.053  | ug/l  | 99       |
| 74) N-amyl acetate             | 10.88 | 43   | 234745   | 41.583  | ug/l  | 96       |
| 75) 1,1,2,2-Tetrachloroethane  | 11.25 | 83   | 187774   | 61.047  | ug/l  | 99       |
| 76) 1,2,3-Trichloropropane     | 11.28 | 75   | 246697   | 63.362  | ug/l  | 84       |
| 77) Bromobenzene               | 11.24 | 156  | 163183   | 50.694  | ug/l  | 94       |
| 78) n-propylbenzene            | 11.34 | 91   | 683007   | 49.812  | ug/l  | 99       |
| 79) 2-Chlorotoluene            | 11.40 | 91   | 415905   | 49.693  | ug/l  | 99       |
| 80) 1,3,5-Trimethylbenzene     | 11.49 | 105  | 526524   | 51.379  | ug/l  | 99       |
| 81) trans-1,4-Dichloro-2-buten | 11.06 | 75   | 70589    | 44.598  | ug/l  | 98       |
| 82) 4-Chlorotoluene            | 11.49 | 91   | 494091   | 49.991  | ug/l  | 98       |
| 83) tert-Butylbenzene          | 11.76 | 119  | 495747   | 56.616  | ug/l  | 95       |
| 84) 1,2,4-Trimethylbenzene     | 11.79 | 105  | 535136   | 51.635  | ug/l  | 99       |
| 85) sec-Butylbenzene           | 11.93 | 105  | 576014   | 52.439  | ug/l  | 99       |
| 86) p-Isopropyltoluene         | 12.05 | 119  | 556119   | 54.047  | ug/l  | 99       |
| 87) 1,3-Dichlorobenzene        | 12.01 | 146  | 295460   | 52.980  | ug/l  | 99       |
| 88) 1,4-Dichlorobenzene        | 12.08 | 146  | 290007   | 50.845  | ug/l  | 97       |
| 89) n-Butylbenzene             | 12.37 | 91   | 457194   | 53.345  | ug/l  | 99       |
| 90) Hexachloroethane           | 12.58 | 117  | 98800    | 55.590  | ug/l  | 97       |
| 91) 1,2-Dichlorobenzene        | 12.38 | 146  | 287319   | 53.177  | ug/l  | 99       |
| 92) 1,2-Dibromo-3-Chloropropan | 12.98 | 75   | 51226    | 51.416  | ug/l  | 96       |

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061220\  
Data File : VX016743.D  
Acq On : 12 Jun 2020 20:50  
Operator : JC/SP  
Sample : VSTDCCC050  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 26 Sample Multiplier: 1


Instrument :  
MSVOA\_X  
ClientSampleId :

Quant Time: Jun 13 04:36:37 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X052720W.M  
Quant Title : SW846 8260  
QLast Update : Wed May 27 16:31:05 2020  
Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|-------|------|----------|--------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 13.63 | 180  | 205668   | 57.315 | ug/l  | 97       |
| 94) Hexachlorobutadiene    | 13.77 | 225  | 84919    | 60.274 | ug/l  | 97       |
| 95) Naphthalene            | 13.82 | 128  | 661051   | 53.738 | ug/l  | 99       |
| 96) 1,2,3-Trichlorobenzene | 14.01 | 180  | 202981   | 57.738 | ug/l  | 98       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X


ClientSampleId :

Tot Ion:168 Resp: 228497

Ion Ratio Lower Upper

168 100

99 56.2 45.2 67.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63

80000

60000



40000

20000

0

5.50 5.60 5.70 5.80

Time-->



#2

Dichlorodifluoromethane

Concen: 43.099 ug/l

RT: 1.18 min Scan# 15

Delta R.T. -0.00 min

Lab File: VX016743.D


Acq: 12 Jun 2020 20:50

Tgt Ion: 85 Resp: 96914

Ion Ratio Lower Upper

85 100

87 32.2 15.4 46.2



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.18

120000

100000

80000

60000


40000


20000

0

1.15 1.20 1.25

Time-->





#3

Chloromethane

Concen: 37.475 ug/l

RT: 1.31 min Scan# 37

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X





ClientSampleId :

Tot Ion: 50 Resp: 101834

Ion Ratio Lower Upper

50 100

52 32.9 26.2 39.2



#4

Vinyl Chloride

Concen: 42.789 ug/l

RT: 1.39 min Scan# 50

Delta R.T. -0.00 min

Lab File: VX016743.D




Acq: 12 Jun 2020 20:50


Tgt Ion: 62 Resp: 122966

Ion Ratio Lower Upper

62 100

64 32.3 25.9 38.9





#5

Bromomethane

Concen: 47.932 ug/l

RT: 1.62 min Scan# 87

Delta R.T. -0.01 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X





ClientSampleId :

Tot Ion: 94 Resp: 66974

Ion Ratio Lower Upper

94 100

96 92.6 75.1 112.7



#6

Chloroethane

Concen: 47.664 ug/l

RT: 1.70 min Scan# 100

Delta R.T. -0.01 min

Lab File: VX016743.D




Acq: 12 Jun 2020 20:50


Tgt Ion: 64 Resp: 79901

Ion Ratio Lower Upper

64 100

66 32.3 26.7 40.1









#7

Trichlorofluoromethane  
Concen: 54.962 ug/l  
RT: 1.91 min Scan# 135  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Instrument :  
MSVOA\_X  
ClientSampleId :




Tot Ion:101 Resp: 207430  
Ion Ratio Lower Upper  
101 100  
103 66.1 50.9 76.3




#8

Diethyl Ether  
Concen: 49.633 ug/l  
RT: 2.17 min Scan# 177  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 74 Resp: 78107  
Ion Ratio Lower Upper  
74 100  
45 88.0 47.5 142.5







#9

1,1,2-Trichlorotrifluoroethane

Concen: 53.722 ug/l

RT: 2.36 min Scan# 209

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion:101 Resp: 111586

Ion Ratio Lower Upper

101 100

85 41.5 34.4 51.6




151 83.9 63.8 95.6



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 45.610 ug/l

RT: 2.49 min Scan# 230

Delta R.T. 0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion:142 Resp: 112477

Ion Ratio Lower Upper

142 100

127 43.6 35.3 52.9



141 14.6 11.4 17.2




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11  
Tert butyl alcohol  
Concen: 232.924 ug/l  
RT: 3.03 min Scan# 318  
Delta R.T. 0.01 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampleId :


Tot Ion: 59 Resp: 169489  
Ion Ratio Lower Upper  
59 100  
57 10.4 8.2 12.4



#12  
1,1-Dichloroethene  
Concen: 50.415 ug/l  
RT: 2.35 min Scan# 207  
Delta R.T. -0.01 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 96 Resp: 110845  
Ion Ratio Lower Upper  
96 100  
61 154.9 126.6 189.8  
98 63.3 51.3 76.9





#13

Acrolein

Concen: 258.916 ug/l

RT: 2.27 min Scan# 194

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50





Instrument :  
MSVOA\_X  
ClientSampleId :

Tot Ion: 56 Resp: 80144

Ion Ratio Lower Upper

56 100

55 71.2 58.3 87.5



#14

Allyl chloride

Concen: 43.099 ug/l

RT: 2.70 min Scan# 265

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50




Tgt Ion: 41 Resp: 174266


Ion Ratio Lower Upper

41 100

39 61.4 47.7 71.5

76 36.8 26.6 40.0





#15

Acrylonitrile

Concen: 239.366 ug/l

RT: 3.12 min Scan# 333

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion: 53 Resp: 356790

Ion Ratio Lower Upper

53 100

52 82.7 66.7 100.1




51 36.3 28.6 43.0



Abundance Ion 53.00 (52.70 to 53.70): VX0

Ion 52.00 (51.70 to 52.70): VX0

Ion 51.00 (50.70 to 51.70): VX0



#16

Acetone

Concen: 241.589 ug/l

RT: 2.42 min Scan# 218

Delta R.T. -0.00 min

Lab File: VX016743.D


Acq: 12 Jun 2020 20:50

Tgt Ion: 43 Resp: 300352

Ion Ratio Lower Upper



43 100


58 33.1 26.1 39.1



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#17  
Carbon Disulfide  
Concen: 42.049 ug/l  
RT: 2.54 min Scan# 239  
Delta R.T. -0.01 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampleId :


Tot Ion: 76 Resp: 274546  
Ion Ratio Lower Upper  
76 100  
78 8.3 7.4 11.0



#18  
Methyl Acetate  
Concen: 55.316 ug/l  
RT: 2.75 min Scan# 272  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 43 Resp: 176385  
Ion Ratio Lower Upper  
43 100  
74 27.4 19.7 29.5









#19

Methyl tert-butyl Ether  
 Concen: 51.340 ug/l  
 RT: 3.17 min Scan# 341  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Instrument :  
 MSVOA\_X  
 ClientSampleId :




Tot Ion: 73 Resp: 405327  
 Ion Ratio Lower Upper  
 73 100  
 57 20.4 17.8 26.6




#20

Methylene Chloride  
 Concen: 48.287 ug/l  
 RT: 2.83 min Scan# 286  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Tgt Ion: 84 Resp: 124204  
 Ion Ratio Lower Upper  
 84 100  
 49 114.1 94.5 141.7  
 51 36.5 30.2 45.4  
 86 65.3 51.1 76.7





#21

trans-1,2-Dichloroethene

Concen: 48.567 ug/l

RT: 3.14 min Scan# 337

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion: 96 Resp: 118111

Ion Ratio Lower Upper

96 100

61 129.8 115.3 172.9




98 62.0 51.9 77.9



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 44.955 ug/l

RT: 3.82 min Scan# 449

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Tgt Ion: 45 Resp: 349193


Ion Ratio Lower Upper

45 100

43 68.6 57.0 85.6

87 31.9 22.5 33.7

59 12.2 9.1 13.7






Abundance Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 221.053 ug/l

RT: 3.78 min Scan# 442

Delta R.T. -0.01 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X


ClientSampleId :

Tot Ion: 43 Resp: 1502925

Ion Ratio Lower Upper




43 100

86 12.1 8.6 13.0



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 47.904 ug/l

RT: 3.67 min Scan# 423

Delta R.T. -0.01 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion: 63 Resp: 210015

Ion Ratio Lower Upper

63 100

98 7.4 3.6 10.9



100 4.8 2.4 7.1




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0







#25

2-Butanone

Concen: 227.703 ug/l

RT: 4.64 min Scan# 582

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X





ClientSampleId :

Tot Ion: 43 Resp: 479682

Ion Ratio Lower Upper

43 100

72 28.3 21.0 31.6



#26

2,2-Dichloropropane

Concen: 46.939 ug/l

RT: 4.54 min Scan# 567

Delta R.T. -0.01 min

Lab File: VX016743.D




Acq: 12 Jun 2020 20:50


Tgt Ion: 77 Resp: 168392

Ion Ratio Lower Upper

77 100

97 23.6 11.5 34.4









#27

cis-1,2-Dichloroethene  
Concen: 48.916 ug/l  
RT: 4.56 min Scan# 570  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Instrument :  
MSVOA\_X  
ClientSampleId :




| Tot Ion | 96    | Resp  | 137622 |
|---------|-------|-------|--------|
| Ion     | Ratio | Lower | Upper  |
| 96      | 100   |       |        |
| 61      | 136.8 | 0.0   | 284.6  |
| 98      | 64.9  | 0.0   | 127.2  |




#28

Bromochloromethane  
Concen: 40.839 ug/l  
RT: 4.98 min Scan# 639  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

| Tgt Ion | 49    | Resp  | 83202 |
|---------|-------|-------|-------|
| Ion     | Ratio | Lower | Upper |
| 49      | 100   |       |       |
| 129     | 2.6   | 0.0   | 4.0   |
| 130     | 95.1  | 63.0  | 94.4# |









#29  
Tetrahydrofuran  
Concen: 226.794 ug/l  
RT: 5.09 min Scan# 657  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampleId :


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 42      | 100   |       |       |
| 72      | 51.1  | 37.0  | 55.4  |
| 71      | 46.9  | 34.1  | 51.1  |



#30  
Chloroform  
Concen: 51.692 ug/l  
RT: 5.18 min Scan# 671  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 83      | 100   |       |       |
| 85      | 63.3  | 52.0  | 78.0  |









#31  
Cyclohexane  
Concen: 42.652 ug/l  
RT: 5.55 min Scan# 732  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampled :


Tot Ion: 56 Resp: 168224  
Ion Ratio Lower Upper  
56 100  
69 33.9 25.7 38.5  
84 90.3 70.0 105.0



#32  
1,1,1-Trichloroethane  
Concen: 53.020 ug/l  
RT: 5.46 min Scan# 718  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 97 Resp: 212042  
Ion Ratio Lower Upper  
97 100  
99 64.2 51.4 77.2  
61 39.2 33.5 50.3









#33  
 1,2-Dichloroethane-d4  
 Concen: 47.617 ug/l  
 RT: 6.03 min Scan# 811  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :


Tot Ion: 65 Resp: 135121  
 Ion Ratio Lower Upper  
 65 100  
 67 51.5 0.0 104.8



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 6.83 min Scan# 942  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Tgt Ion: 114 Resp: 342987  
 Ion Ratio Lower Upper  
 114 100  
 63 19.1 0.0 42.6  
 88 16.7 0.0 32.8





#35

Dibromofluoromethane

Concen: 53.116 ug/l

RT: 5.47 min Scan# 719

Delta R.T. 0.01 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampled :


Tot Ion:113 Resp: 112178

Ion Ratio Lower Upper

113 100

111 101.7 83.0 124.6




192 23.0 17.7 26.5



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 49.564 ug/l

RT: 5.77 min Scan# 768

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion: 75 Resp: 160955

Ion Ratio Lower Upper

75 100

110 37.0 17.4 52.3



77 31.4 24.2 36.4




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37  
Ethyl Acetate  
Concen: 44.754 ug/l  
RT: 4.79 min Scan# 608  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampled :


Tot Ion: 43 Resp: 174486  
Ion Ratio Lower Upper  
43 100  
61 15.0 11.3 16.9  
70 13.0 9.3 13.9



#38  
Carbon Tetrachloride  
Concen: 56.089 ug/l  
RT: 5.75 min Scan# 765  
Delta R.T. -0.01 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 117 Resp: 191415  
Ion Ratio Lower Upper  
117 100  
119 92.3 76.3 114.5  
121 29.4 24.6 37.0









#39  
Methylvlcyclohexane  
Concen: 49.698 ug/l  
RT: 7.44 min Scan# 1042  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampleId :

Tot Ion: 83 Resp: 175299  
Ion Ratio Lower Upper  
83 100  
55 77.6 61.7 92.5  
98 49.7 38.7 58.1




#40  
Benzene  
Concen: 50.112 ug/l  
RT: 6.11 min Scan# 824  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 78 Resp: 484622  
Ion Ratio Lower Upper  
78 100  
77 24.5 18.9 28.3











#41  
Methacrylonitrile  
Concen: 45.423 ug/l  
RT: 5.01 min Scan# 643  
Delta R.T. 0.01 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampled :


|          |       |       |       |
|----------|-------|-------|-------|
| Tot Ion: | 41    | Resp: | 97276 |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 54.2  | 43.0  | 64.4  |
| 67       | 80.5  | 58.9  | 88.3  |
| 52       | 31.3  | 23.1  | 34.7  |



#42  
1,2-Dichloroethane  
Concen: 51.877 ug/l  
RT: 6.16 min Scan# 832  
Delta R.T. -0.01 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 62    | Resp: | 178412 |
| Ion      | Ratio | Lower | Upper  |
| 62       | 100   |       |        |
| 98       | 10.0  | 0.0   | 19.8   |





#43

Isopropyl Acetate

Concen: 46.459 ug/l

RT: 6.41 min Scan# 873

Delta R.T. -0.01 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampled :


Tot Ion: 43 Resp: 282538

Ion Ratio Lower Upper

43 100

61 22.5 16.8 25.2




87 15.6 11.2 16.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#44

Trichloroethene

Concen: 48.372 ug/l

RT: 7.19 min Scan# 1001

Delta R.T. -0.00 min

Lab File: VX016743.D


Acq: 12 Jun 2020 20:50

Tgt Ion:130 Resp: 148419

Ion Ratio Lower Upper



130 100


95 94.3 0.0 198.2



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0





#45

1,2-Dichloropropane

Concen: 49.668 ug/l

RT: 7.49 min Scan# 1050

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X





ClientSampleId :

Tot Ion: 63 Resp: 122279

Ion Ratio Lower Upper

63 100

65 30.9 24.2 36.2



#46

Dibromomethane

Concen: 52.889 ug/l

RT: 7.64 min Scan# 1075

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50




Tgt Ion: 93 Resp: 87458


Ion Ratio Lower Upper

93 100

95 83.4 66.6 100.0

174 114.9 86.2 129.4





#47

Bromodichloromethane

Concen: 55.084 ug/l

RT: 7.88 min Scan# 1114

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion: 83 Resp: 188603

Ion Ratio Lower Upper

83 100

85 65.0 52.0 78.0




127 9.1 7.4 11.0



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 47.650 ug/l

RT: 7.74 min Scan# 1092

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion: 41 Resp: 138749

Ion Ratio Lower Upper

41 100

69 94.1 68.6 102.8



39 54.3 42.4 63.6




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: 931.606 ug/l

RT: 7.71 min Scan# 1087

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Instrument :

MSVOA\_X

ClientSampleId :

Tot Ion: 88 Resp: 63483


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 28.8  | 24.2  | 36.4  |
| 58  | 68.2  | 56.5  | 84.7  |





Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



Time-->



#50

Toluene-d8

Concen: 50.189 ug/l

RT: 8.70 min Scan# 1248


Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion: 98 Resp: 412560

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 98  | 100   |       |       |
| 100 | 68.1  | 53.8  | 80.6  |





Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX0



Time-->









#51  
 4-Methyl-2-Pentanone  
 Concen: 244.232 ug/l  
 RT: 8.62 min Scan# 1236  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :


Tot Ion: 43 Resp: 924833  
 Ion Ratio Lower Upper  
 43 100  
 58 43.2 33.0 49.6



#52  
 Toluene  
 Concen: 52.280 ug/l  
 RT: 8.77 min Scan# 1260  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Tgt Ion: 92 Resp: 318820  
 Ion Ratio Lower Upper  
 92 100  
 91 168.3 135.0 202.4









#53  
 t-1,3-Dichloropropene  
 Concen: 53.115 ug/l  
 RT: 9.02 min Scan# 1302  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampled :


Tot Ion: 75 Resp: 206798  
 Ion Ratio Lower Upper  
 75 100  
 77 31.5 25.5 38.3



#54  
 cis-1,3-Dichloropropene  
 Concen: 51.818 ug/l  
 RT: 8.42 min Scan# 1202  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Tgt Ion: 75 Resp: 213542  
 Ion Ratio Lower Upper  
 75 100  
 77 31.8 24.6 36.8  
 39 43.1 35.3 52.9









#55  
 1,1,2-Trichloroethane  
 Concen: 55.153 ug/l  
 RT: 9.20 min Scan# 1330  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Instrument :  
 MSVOA\_X  
 ClientSampleId :

|          |       |       |        |
|----------|-------|-------|--------|
| Tot Ion: | 97    | Resp: | 133729 |
| Ion      | Ratio | Lower | Upper  |
| 97       | 100   |       |        |
| 83       | 82.5  | 70.9  | 106.3  |
| 85       | 52.3  | 43.4  | 65.2   |
| 99       | 62.8  | 51.0  | 76.6   |




Abundance Ion 96.90 (96.60 to 97.60): VX0  
 Ion 82.90 (82.60 to 83.60): VX0  
 Ion 84.90 (84.60 to 85.60): VX0  
 Ion 98.90 (98.60 to 99.60): VX0





#56  
 Ethyl methacrylate  
 Concen: 52.127 ug/l  
 RT: 9.16 min Scan# 1324  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 69    | Resp: | 205845 |
| Ion      | Ratio | Lower | Upper  |
| 69       | 100   |       |        |
| 41       | 59.3  | 50.7  | 76.1   |
| 39       | 31.4  | 25.8  | 38.6   |



Abundance Ion 69.00 (68.70 to 69.70): VX0  
 Ion 41.00 (40.70 to 41.70): VX0  
 Ion 39.00 (38.70 to 39.70): VX0











#57  
 1,3-Dichloropropane  
 Concen: 52.772 ug/l  
 RT: 9.35 min Scan# 1356  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :


Tot Ion: 76 Resp: 219628  
 Ion Ratio Lower Upper  
 76 100  
 78 32.6 26.2 39.4



#58  
 2-Chloroethyl Vinyl ether  
 Concen: 237.239 ug/l  
 RT: 8.29 min Scan# 1182  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Tgt Ion: 63 Resp: 488990  
 Ion Ratio Lower Upper  
 63 100  
 106 31.9 23.2 34.8









#59  
2-Hexanone  
Concen: 244.803 ug/l  
RT: 9.48 min Scan# 1376  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampleId :


Tot Ion: 43 Resp: 720883  
Ion Ratio Lower Upper  
43 100  
58 60.2 28.8 86.4



#60  
Dibromochloromethane  
Concen: 58.780 ug/l  
RT: 9.57 min Scan# 1391  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 129 Resp: 160390  
Ion Ratio Lower Upper  
129 100  
127 77.0 38.6 116.0





#61

1,2-Dibromoethane

Concen: 55.985 ug/l

RT: 9.65 min Scan# 1405

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X


ClientSampleId :

Tot Ion:107 Resp: 145363

Ion Ratio Lower Upper




107 100

109 92.6 75.5 113.3



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 51.715 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion: 95 Resp: 164257

Ion Ratio Lower Upper

95 100

174 85.7 0.0 163.0



176 83.2 0.0 156.8




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.10 min Scan# 1478  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampleId :


Tot Ion:117 Resp: 328203  
Ion Ratio Lower Upper  
117 100  
82 54.3 45.0 67.6  
119 32.0 25.8 38.8



#64  
Tetrachloroethene  
Concen: 41.464 ug/l  
RT: 9.32 min Scan# 1350  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion:164 Resp: 152252  
Ion Ratio Lower Upper  
164 100  
166 129.9 100.9 151.3  
129 97.3 73.2 109.8  
131 90.3 70.8 106.2





#65

Chlorobenzene

Concen: 52.972 ug/l

RT: 10.12 min Scan# 1482

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X





ClientSampleId :

Tot Ion:112 Resp: 354446

Ion Ratio Lower Upper

112 100

114 32.5 26.0 39.0



#66

1,1,1,2-Tetrachloroethane

Concen: 56.310 ug/l

RT: 10.21 min Scan# 1496

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50




Tgt Ion:131 Resp: 143827


Ion Ratio Lower Upper

131 100

133 94.7 48.4 145.3

119 62.2 32.0 96.2









#67  
Ethyl Benzene  
Concen: 51.490 ug/l  
RT: 10.24 min Scan# 1501  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampled :


Tot Ion: 91 Resp: 613407  
Ion Ratio Lower Upper  
91 100  
106 31.3 25.0 37.4



#68  
m/p-Xylenes  
Concen: 106.237 ug/l  
RT: 10.34 min Scan# 1518  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion: 106 Resp: 469251  
Ion Ratio Lower Upper  
106 100  
91 204.0 163.8 245.8









#69  
o-Xylene  
Concen: 53.565 ug/l  
RT: 10.68 min Scan# 1574  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50




Instrument :  
MSVOA\_X  
ClientSampled :


Tot Ion:106 Resp: 229136  
Ion Ratio Lower Upper  
106 100  
91 214.2 109.6 328.8



#70  
Styrene  
Concen: 55.632 ug/l  
RT: 10.70 min Scan# 1576  
Delta R.T. -0.00 min  
Lab File: VX016743.D  
Acq: 12 Jun 2020 20:50

Tgt Ion:104 Resp: 409050  
Ion Ratio Lower Upper  
104 100  
78 49.0 40.2 60.2  
103 53.8 43.1 64.7





#71

Bromoform

Concen: 60.414 ug/l

RT: 10.84 min Scan# 1600

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampled :


Tot Ion:173 Resp: 130412

Ion Ratio Lower Upper


173 100

175 49.0 23.9 71.7



254 0.1 0.1 0.1



Abundance Ion 172.70 (172.40 to 173.40): V  
Ion 174.70 (174.40 to 175.40): V  
Ion 254.40 (254.10 to 255.10): V



Time-->



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.06 min Scan# 1800

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion:152 Resp: 177131

Ion Ratio Lower Upper


152 100

115 80.8 39.9 119.6


150 170.5 0.0 341.0




Abundance Ion 151.90 (151.60 to 152.60): V  
Ion 114.90 (114.60 to 115.60): V  
Ion 149.90 (149.60 to 150.60): V



Time-->







#73

Isopropylbenzene

Concen: 50.053 ug/l

RT: 11.00 min Scan# 1626

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X


ClientSampleId :

Tot Ion:105 Resp: 614309

Ion Ratio Lower Upper




105 100

120 26.5 13.1 39.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 41.583 ug/l

RT: 10.88 min Scan# 1606

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Tgt Ion: 43 Resp: 234745


Ion Ratio Lower Upper

43 100

70 48.8 36.1 54.1

55 27.0 20.2 30.4

61 27.0 20.3 30.5






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0





#75

1,1,2,2-Tetrachloroethane

Concen: 61.047 ug/l

RT: 11.25 min Scan# 1667

Delta R.T. -0.01 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion: 83 Resp: 187774

Ion Ratio Lower Upper

83 100

131 11.1 5.1 15.3




85 64.6 32.0 96.2



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): VX0

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 63.362 ug/l

RT: 11.28 min Scan# 1672

Delta R.T. -0.00 min

Lab File: VX016743.D


Acq: 12 Jun 2020 20:50

Tgt Ion: 75 Resp: 246697

Ion Ratio Lower Upper



75 100


77 32.4 21.3 63.7



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 50.694 ug/l

RT: 11.24 min Scan# 1665

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion:156 Resp: 163183

Ion Ratio Lower Upper

156 100

77 149.2 79.6 238.8




158 99.1 48.1 144.4



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 49.812 ug/l

RT: 11.34 min Scan# 1682

Delta R.T. -0.00 min

Lab File: VX016743.D


Acq: 12 Jun 2020 20:50

Tgt Ion: 91 Resp: 683007

Ion Ratio Lower Upper



91 100


120 23.7 11.6 34.7



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V









#79  
 2-Chlorotoluene  
 Concen: 49.693 ug/l  
 RT: 11.40 min Scan# 1692  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :


Tot Ion: 91 Resp: 415905  
 Ion Ratio Lower Upper  
 91 100  
 126 33.9 16.8 50.3



#80  
 1,3,5-Trimethylbenzene  
 Concen: 51.379 ug/l  
 RT: 11.49 min Scan# 1707  
 Delta R.T. 0.01 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

Tgt Ion: 105 Resp: 526524  
 Ion Ratio Lower Upper  
 105 100  
 120 49.0 24.3 72.8





#81

trans-1,4-Dichloro-2-butene

Concen: 44.598 ug/l

RT: 11.06 min Scan# 1636

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion: 75 Resp: 70589

Ion Ratio Lower Upper

75 100

53 88.3 71.3 106.9




89 48.9 36.6 55.0



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#82

4-Chlorotoluene

Concen: 49.991 ug/l

RT: 11.49 min Scan# 1707

Delta R.T. -0.00 min

Lab File: VX016743.D


Acq: 12 Jun 2020 20:50

Tgt Ion: 91 Resp: 494091

Ion Ratio Lower Upper



91 100


126 30.2 14.7 44.1



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V









#83  
 tert-Butylbenzene  
 Concen: 56.616 ug/l  
 RT: 11.76 min Scan# 1750  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 119     | 100   |       |       |
| 91      | 57.3  | 31.3  | 93.8  |
| 134     | 23.2  | 12.0  | 36.0  |



#84  
 1,2,4-Trimethylbenzene  
 Concen: 51.635 ug/l  
 RT: 11.79 min Scan# 1756  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 105     | 100   |       |       |
| 120     | 45.6  | 22.6  | 67.7  |





#85

sec-Butylbenzene

Concen: 52.439 ug/l

RT: 11.93 min Scan# 1779

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X


ClientSampled :

Tot Ion:105 Resp: 576014

Ion Ratio Lower Upper




105 100

134 20.8 10.1 30.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 54.047 ug/l

RT: 12.05 min Scan# 1798

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion:119 Resp: 556119

Ion Ratio Lower Upper

119 100

134 25.6 12.7 38.0



91 23.1 11.7 35.0




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V









#87  
 1,3-Dichlorobenzene  
 Concen: 52.980 ug/l  
 RT: 12.01 min Scan# 1792  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 111     | 39.3  | 20.0  | 60.0  |
| 148     | 63.0  | 31.9  | 95.5  |




#88  
 1,4-Dichlorobenzene  
 Concen: 50.845 ug/l  
 RT: 12.08 min Scan# 1803  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 111     | 39.1  | 19.4  | 58.2  |
| 148     | 66.2  | 31.4  | 94.3  |







#89

n-Butylbenzene

Concen: 53.345 ug/l

RT: 12.37 min Scan# 1851

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :





Tot Ion: 91 Resp: 457194

Ion Ratio Lower Upper

91 100

92 54.4 27.2 81.5

134 28.1 13.1 39.3



#90

Hexachloroethane

Concen: 55.590 ug/l

RT: 12.58 min Scan# 1885

Delta R.T. -0.00 min

Lab File: VX016743.D




Acq: 12 Jun 2020 20:50


Tgt Ion: 117 Resp: 98800

Ion Ratio Lower Upper

117 100

201 89.2 43.1 129.3





#91

1,2-Dichlorobenzene

Concen: 53.177 ug/l

RT: 12.38 min Scan# 1852

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion:146 Resp: 287319

Ion Ratio Lower Upper

146 100

111 40.8 21.4 64.3




148 63.7 31.9 95.7



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 51.416 ug/l

RT: 12.98 min Scan# 1951

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion: 75 Resp: 51226

Ion Ratio Lower Upper

75 100

155 94.4 44.1 132.4



157 118.1 57.8 173.3




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93  
 1,2,4-Trichlorobenzene  
 Concen: 57.315 ug/l  
 RT: 13.63 min Scan# 2057  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50




Instrument :  
 MSVOA\_X  
 ClientSampleId :


| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 180     | 100   |       |       |
| 182     | 92.6  | 48.3  | 144.8 |
| 145     | 30.5  | 15.4  | 46.2  |



#94  
 Hexachlorobutadiene  
 Concen: 60.274 ug/l  
 RT: 13.77 min Scan# 2080  
 Delta R.T. -0.00 min  
 Lab File: VX016743.D  
 Acq: 12 Jun 2020 20:50

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 225     | 100   |       |       |
| 223     | 60.9  | 32.3  | 96.8  |
| 227     | 62.8  | 31.6  | 94.7  |





#95

Naphthalene

Concen: 53.738 ug/l

RT: 13.82 min Scan# 2088

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion:128 Resp: 661051

Ion Ratio Lower Upper

128 100

127 12.9 10.6 15.8




129 11.1 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 57.738 ug/l

RT: 14.01 min Scan# 2119

Delta R.T. -0.00 min

Lab File: VX016743.D

Acq: 12 Jun 2020 20:50


Tgt Ion:180 Resp: 202981

Ion Ratio Lower Upper

180 100

182 95.9 48.6 145.8

145 31.7 16.6 49.8



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

