

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX061220\
 Data File : VX016737.D
 Acq On : 12 Jun 2020 18:34
 Operator : JC/SP
 Sample : L2971-01 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30698

Quant Time: Jun 15 08:17:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	216238	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	341362	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	350951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	184425	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	130820	48.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.44%	
35) Dibromofluoromethane	5.46	113	110899	52.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.52%	
50) Toluene-d8	8.70	98	423196	51.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.46%	
62) 4-Bromofluorobenzene	11.12	95	167656	53.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.08%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX061220\
Data File : VX016737.D
Acq On : 12 Jun 2020 18:34
Operator : JC/SP
Sample : L2971-01 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30698

Quant Time: Jun 15 08:17:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

