

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX061220\
 Data File : VX016738.D
 Acq On : 12 Jun 2020 18:57
 Operator : JC/SP
 Sample : L2971-03 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40755

Quant Time: Jun 15 08:20:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	219287	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	346992	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	365947	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	193804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	132939	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
35) Dibromofluoromethane	5.46	113	109393	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	8.70	98	430122	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.12	95	175028	54.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.94%	
Target Compounds						
16) Acetone	2.42	43	32659	27.373	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX061220\
Data File : VX016738.D
Acq On : 12 Jun 2020 18:57
Operator : JC/SP
Sample : L2971-03 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40755

Quant Time: Jun 15 08:20:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

